

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE092616\
 Data File : BE092430.D
 Acq On : 26 Sep 2016 17:48
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 26 19:24:45 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE092216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 22 18:25:56 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	154	-0.02
2	1,4-Dioxane	0.400	0.329	17.8	140	-0.02
3	n-Nitrosodimethylamine	0.400	0.478	-19.5	172	-0.02
4 S	2-Fluorophenol	0.400	0.403	-0.8	144	-0.01
5 S	Phenol-d6	0.400	0.346	13.5	142	0.00
6	bis(2-Chloroethyl)ether	0.400	0.412	-3.0	175	-0.02
7 I	Naphthalene-d8	5.000	5.000	0.0	150	0.00
8 S	Nitrobenzene-d5	0.400	0.441	-10.2	162	0.00
9	Naphthalene	0.400	0.385	3.8	145	-0.02
10	Hexachlorobutadiene	0.400	0.375	6.3	142	-0.02
11	2-Methylnaphthalene	0.400	0.380	5.0	145	-0.01
12 I	Acenaphthene-d10	5.000	5.000	0.0	142	0.00
13 S	2,4,6-Tribromophenol	0.400	0.412	-3.0	130	0.00
14 S	2-Fluorobiphenyl	0.400	0.405	-1.3	146	0.00
15	Acenaphthylene	0.400	0.369	7.8	135	0.00
16	Acenaphthene	0.400	0.389	2.8	134	0.00
17	Fluorene	0.400	0.373	6.8	134	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	131	0.02
19	Hexachlorobenzene	0.400	0.409	-2.2	143	0.00
20	Pentachlorophenol	0.400	0.393	1.8	135	0.00
21	Phenanthrene	0.400	0.404	-1.0	139	0.00
22	Anthracene	0.400	0.391	2.3	140	0.02
23	Fluoranthene	0.400	0.417	-4.2	148	0.00
24 I	Chrysene-d12	5.000	5.000	0.0	133	0.00
25	Pyrene	0.400	0.444	-11.0	153	0.00
26 S	Terphenyl-d14	0.400	0.424	-6.0	142	0.00
27	Benzo(a)anthracene	0.400	0.430	-7.5	146	0.00
28	Chrysene	0.400	0.402	-0.5	124	0.00
29	Bis(2-ethylhexyl)phthalate	0.400	0.408	-2.0	134	0.00
30	Indeno(1,2,3-cd)pyrene	0.400	0.392	2.0	132	0.03
31 I	Perylene-d12	5.000	5.000	0.0	136	0.04
32	Benzo(b)fluoranthene	0.400	0.380	5.0	130	0.02
33	Benzo(k)fluoranthene	0.400	0.398	0.5	145	0.04
34 C	Benzo(a)pyrene	0.400	0.390	2.5	140	0.02
35	Dibenzo(a,h)anthracene	0.400	0.382	4.5	136	0.05
36	Benzo(g,h,i)perylene	0.400	0.372	7.0	130	0.05

(#) = Out of Range

SPCC's out = 0 CCC's out = 0