

Data Path : Z:\HPCHEM1\BNA E\Data\BE092617\
 Data File : BE094123.D
 Acq On : 26 Sep 2017 9:09
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 26 15:48:44 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE092517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 25 17:37:08 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	112	0.00
2	1,4-Dioxane	0.780	0.730	6.4	101	0.00
3	n-Nitrosodimethylamine	0.996	0.791	20.6	90	0.00
4 S	2-Fluorophenol	1.516	1.404	7.4	103	0.00
5 S	Phenol-d6	2.106	2.151	-2.1	114	0.00
6	bis(2-Chloroethyl)ether	1.632	1.591	2.5	107	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	125	0.00
8 S	Nitrobenzene-d5	0.387	0.336	13.2	115	0.00
9	Naphthalene	4.516	4.483	0.7	125	0.00
10	Hexachlorobutadiene	0.169	0.161	4.7	124	0.00
11 SURR	2-Methylnaphthalene-d10	0.616	0.646	-4.9	133	0.00
12	2-Methylnaphthalene	0.731	0.768	-5.1	135	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	142	0.00
14 S	2,4,6-Tribromophenol	0.256	0.193	24.6	125	0.00
15 S	2-Fluorobiphenyl	1.322	1.289	2.5	132	0.00
16	Acenaphthylene	2.005	2.034	-1.4	140	0.00
17	Acenaphthene	1.253	1.348	-7.6	142	0.00
18	Fluorene	1.700	1.789	-5.2	144	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	127	0.00
20	4-Bromophenyl-phenylether	0.164	0.193	-17.7	153#	0.00
21	Hexachlorobenzene	0.104	0.144	-38.5#	180#	0.00
22	Pentachlorophenol	0.095	0.070	26.3#	96	0.00
23	Phenanthrene	1.023	1.184	-15.7	153#	0.00
24	Anthracene	1.074	1.054	1.9	132	0.00
25 SURR	Fluoranthene-d10	3.873	3.611	6.8	125	0.00
26	Fluoranthene	1.147	1.083	5.6	124	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	107	0.00
28	Pvrene	1.267	1.384	-9.2	121	0.00
29 S	Terphenyl-d14	0.730	0.766	-4.9	118	0.00
30	Benzo(a)anthracene	1.347	1.320	2.0	105	0.00
31	Chrysene	1.300	1.281	1.5	106	0.00
32	Bis(2-ethylhexyl)phthalate	6.447	3.644	43.5#	81	0.00
33	Indeno(1,2,3-cd)pyrene	1.391	1.308	6.0	105	0.00
34 I	Perylene-d12	1.000	1.000	0.0	104	0.00
35	Benzo(b)fluoranthene	1.443	1.327	8.0	107	0.00
36	Benzo(k)fluoranthene	1.382	1.285	7.0	97	0.00
37 C	Benzo(a)pyrene	1.304	1.254	3.8	104	0.00
38	Dibenzo(a,h)anthracene	1.235	1.174	4.9	104	0.00
39	Benzo(a,h,i)perylene	1.222	1.166	4.6	104	0.00

Data Path : Z:\HPCHEM1\BNA E\Data\BE092617\
Data File : BE094123.D
Acq On : 26 Sep 2017 9:09
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 26 15:48:44 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE092517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 25 17:37:08 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			