

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE092618\
 Data File : BE097681.D
 Acq On : 26 Sep 2018 10:23
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 26 15:30:55 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE092418.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 24 18:40:34 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	-0.01
2	1,4-Dioxane	0.716	0.598	16.5	99	0.00
3	n-Nitrosodimethylamine	0.421	0.506	-20.2	135	0.00
4 S	2-Fluorophenol	1.053	1.051	0.2	119	0.00
5 S	Phenol-d6	1.487	1.588	-6.8	128	-0.01
6	bis(2-Chloroethyl)ether	1.204	1.280	-6.3	108	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	105	0.00
8 S	Nitrobenzene-d5	0.232	0.270	-16.4	130	-0.01
9	Naphthalene	3.636	3.732	-2.6	116	-0.01
10	Hexachlorobutadiene	0.165	0.160	3.0	111	0.00
11 SURR	2-Methylnaphthalene-d10	0.592	0.623	-5.2	118	0.00
12	2-Methylnaphthalene	0.576	0.622	-8.0	122	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	107	0.00
14 S	2,4,6-Tribromophenol	0.155	0.157	-1.3	124	0.00
15 S	2-Fluorobiphenyl	1.305	1.354	-3.8	118	-0.01
16	Acenaphthylene	1.608	1.597	0.7	118	0.00
17	Acenaphthene	1.050	1.089	-3.7	119	0.00
18	Fluorene	1.477	1.450	1.8	114	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	96	0.00
20	4-Bromophenyl-phenylether	0.177	0.189	-6.8	110	0.00
21	Hexachlorobenzene	0.204	0.217	-6.4	109	0.00
22	Pentachlorophenol	0.049	0.048	2.0	135	0.00
23	Phenanthrene	0.946	0.973	-2.9	106	0.00
24	Anthracene	0.786	0.805	-2.4	110	0.00
25 SURR	Fluoranthene-d10	4.361	4.199	3.7	103	0.00
26	Fluoranthene	1.135	1.094	3.6	103	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	91	0.00
28	Pvrene	1.044	1.103	-5.7	103	0.00
29 S	Terphenyl-d14	0.787	0.807	-2.5	102	0.00
30	Benzo(a)anthracene	0.945	0.936	1.0	104	0.00
31	Chrysene	1.106	1.184	-7.1	100	0.00
32	Bis(2-ethylhexyl)phthalate	1.080	1.194	-10.6	115	0.00
33	Indeno(1,2,3-cd)pyrene	1.165	1.183	-1.5	98	0.00
34 I	Perylene-d12	1.000	1.000	0.0	91	0.00
35	Benzo(b)fluoranthene	1.023	1.033	-1.0	99	0.00
36	Benzo(k)fluoranthene	1.211	1.251	-3.3	101	0.00
37 C	Benzo(a)pyrene	1.010	1.022	-1.2	102	0.00
38	Dibenzo(a,h)anthracene	1.062	1.058	0.4	101	0.00
39	Benzo(a,h,i)perylene	1.056	1.072	-1.5	100	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE092618\
Data File : BE097681.D
Acq On : 26 Sep 2018 10:23
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 26 15:30:55 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE092418.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 24 18:40:34 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			