

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE092618\
 Data File : BE097688.D
 Acq On : 26 Sep 2018 17:15
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 26 18:01:43 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE092418.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 24 18:40:34 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	-0.01
2	1,4-Dioxane	0.716	0.644	10.1	99	0.00
3	n-Nitrosodimethylamine	0.421	0.571	-35.6#	141	-0.01
4 S	2-Fluorophenol	1.053	1.235	-17.3	130	-0.01
5 S	Phenol-d6	1.487	1.823	-22.6	137	-0.01
6	bis(2-Chloroethyl)ether	1.204	1.410	-17.1	111	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	104	-0.01
8 S	Nitrobenzene-d5	0.232	0.278	-19.8	133	-0.01
9	Naphthalene	3.636	3.722	-2.4	114	-0.01
10	Hexachlorobutadiene	0.165	0.161	2.4	110	-0.01
11 SURR	2-Methylnaphthalene-d10	0.592	0.624	-5.4	117	0.00
12	2-Methylnaphthalene	0.576	0.636	-10.4	123	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	107	0.00
14 S	2,4,6-Tribromophenol	0.155	0.170	-9.7	135	0.00
15 S	2-Fluorobiphenyl	1.305	1.348	-3.3	118	-0.01
16	Acenaphthylene	1.608	1.638	-1.9	122	0.00
17	Acenaphthene	1.050	1.102	-5.0	121	0.00
18	Fluorene	1.477	1.493	-1.1	118	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	99	0.00
20	4-Bromophenyl-phenylether	0.177	0.187	-5.6	112	0.00
21	Hexachlorobenzene	0.204	0.210	-2.9	109	0.00
22	Pentachlorophenol	0.049	0.055	-12.2	160#	-0.01
23	Phenanthrene	0.946	0.968	-2.3	109	0.00
24	Anthracene	0.786	0.825	-5.0	116	0.00
25 SURR	Fluoranthene-d10	4.361	4.258	2.4	107	0.00
26	Fluoranthene	1.135	1.103	2.8	107	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	96	0.00
28	Pvrene	1.044	1.099	-5.3	108	0.00
29 S	Terphenyl-d14	0.787	0.808	-2.7	107	0.00
30	Benzo(a)anthracene	0.945	0.974	-3.1	113	0.00
31	Chrysene	1.106	1.137	-2.8	101	0.00
32	Bis(2-ethylhexyl)phthalate	1.080	1.446	-33.9#	146	0.00
33	Indeno(1,2,3-cd)pyrene	1.165	1.226	-5.2	107	0.00
34 I	Perylene-d12	1.000	1.000	0.0	97	0.00
35	Benzo(b)fluoranthene	1.023	1.048	-2.4	108	0.00
36	Benzo(k)fluoranthene	1.211	1.215	-0.3	105	0.00
37 C	Benzo(a)pyrene	1.010	1.033	-2.3	110	0.00
38	Dibenzo(a,h)anthracene	1.062	1.077	-1.4	111	0.00
39	Benzo(a,h,i)perylene	1.056	1.070	-1.3	107	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE092618\
Data File : BE097688.D
Acq On : 26 Sep 2018 17:15
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 26 18:01:43 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE092418.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 24 18:40:34 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			