

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE092619\  
 Data File : BE100569.D  
 Acq On : 26 Sep 2019 16:00  
 Operator : JU  
 Sample : SSTDCCC0.4  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_E  
 LabSampleId :  
 SSTDCCC0.4

Quant Time: Sep 26 18:27:08 2019  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE092019.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Sep 20 18:26:17 2019  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|         | Compound                   | AvqRF | CCRF  | %Dev    | Area% | Dev(min) |
|---------|----------------------------|-------|-------|---------|-------|----------|
| 1 I     | 1,4-Dichlorobenzene-d4     | 1.000 | 1.000 | 0.0     | 105   | 0.00     |
| 2       | 1,4-Dioxane                | 0.577 | 0.608 | -5.4    | 110   | 0.04     |
| 3       | n-Nitrosodimethylamine     | 0.339 | 0.372 | -9.7    | 152#  | 0.02     |
| 4 S     | 2-Fluorophenol             | 0.936 | 1.258 | -34.4#  | 153#  | 0.02     |
| 5 S     | Phenol-d6                  | 1.282 | 1.655 | -29.1#  | 150#  | 0.00     |
| 6       | bis(2-Chloroethyl)ether    | 1.297 | 1.329 | -2.5    | 114   | 0.00     |
| 7 I     | Naphthalene-d8             | 1.000 | 1.000 | 0.0     | 108   | 0.00     |
| 8 S     | Nitrobenzene-d5            | 0.231 | 0.299 | -29.4#  | 156#  | 0.00     |
| 9       | Naphthalene                | 4.147 | 4.016 | 3.2     | 107   | 0.00     |
| 10      | Hexachlorobutadiene        | 0.133 | 0.126 | 5.3     | 104   | 0.00     |
| 11 SURR | 2-Methylnaphthalene-d10    | 0.598 | 0.565 | 5.5     | 108   | 0.00     |
| 12      | 2-Methylnaphthalene        | 0.660 | 0.645 | 2.3     | 112   | 0.00     |
| 13 I    | Acenaphthene-d10           | 1.000 | 1.000 | 0.0     | 97    | 0.00     |
| 14 S    | 2,4,6-Tribromophenol       | 0.065 | 0.094 | -44.6#  | 156#  | 0.00     |
| 15 S    | 2-Fluorobiphenyl           | 1.428 | 1.441 | -0.9    | 101   | 0.00     |
| 16      | Acenaphthylene             | 1.647 | 1.951 | -18.5   | 127   | 0.01     |
| 17      | Acenaphthene               | 1.112 | 1.138 | -2.3    | 106   | 0.00     |
| 18      | Fluorene                   | 1.391 | 1.402 | -0.8    | 105   | 0.00     |
| 19 I    | Phenanthrene-d10           | 1.000 | 1.000 | 0.0     | 87    | 0.00     |
| 20      | 4-Bromophenyl-phenylether  | 0.188 | 0.198 | -5.3    | 95    | 0.00     |
| 21      | Hexachlorobenzene          | 0.164 | 0.157 | 4.3     | 87    | 0.00     |
| 22      | Pentachlorophenol          | 0.037 | 0.056 | -51.4#  | 222#  | -0.01    |
| 23      | Phenanthrene               | 1.115 | 1.156 | -3.7    | 93    | 0.00     |
| 24      | Anthracene                 | 0.938 | 1.094 | -16.6   | 112   | 0.00     |
| 25 SURR | Fluoranthene-d10           | 4.758 | 4.459 | 6.3     | 86    | 0.00     |
| 26      | Fluoranthene               | 1.344 | 1.275 | 5.1     | 87    | 0.00     |
| 27 I    | Chrysene-d12               | 1.000 | 1.000 | 0.0     | 79    | 0.00     |
| 28      | Pvrene                     | 1.151 | 1.304 | -13.3   | 92    | 0.00     |
| 29 S    | Terphenyl-d14              | 0.947 | 0.998 | -5.4    | 86    | 0.00     |
| 30      | Benzo(a)anthracene         | 1.068 | 1.261 | -18.1   | 98    | 0.01     |
| 31      | Chrysene                   | 1.217 | 1.190 | 2.2     | 81    | 0.00     |
| 32      | Bis(2-ethylhexyl)phthalate | 1.151 | 3.454 | -200.1# | 262#  | 0.00     |
| 33      | Indeno(1,2,3-cd)pyrene     | 1.353 | 1.500 | -10.9   | 90    | 0.00     |
| 34 I    | Perylene-d12               | 1.000 | 1.000 | 0.0     | 83    | 0.00     |
| 35      | Benzo(b)fluoranthene       | 1.082 | 1.101 | -1.8    | 93    | 0.00     |
| 36      | Benzo(k)fluoranthene       | 1.252 | 1.107 | 11.6    | 79    | 0.00     |
| 37 C    | Benzo(a)pyrene             | 1.010 | 1.033 | -2.3    | 91    | 0.00     |
| 38      | Dibenzo(a,h)anthracene     | 1.052 | 1.059 | -0.7    | 91    | 0.00     |
| 39      | Benzo(a,h,i)perylene       | 1.098 | 1.070 | 2.6     | 84    | 0.01     |

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE092619\  
Data File : BE100569.D  
Acq On : 26 Sep 2019 16:00  
Operator : JU  
Sample : SSTDCCC0.4  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
BNA\_E  
LabSampleId :  
SSTDCCC0.4

Quant Time: Sep 26 18:27:08 2019  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE092019.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Fri Sep 20 18:26:17 2019  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 25% Max. Rel. Area : 150%

| Compound           | AvgRF                        | CCRF | %Dev Area | % Dev(min) |
|--------------------|------------------------------|------|-----------|------------|
| -----              |                              |      |           |            |
| (#) = Out of Range | SPCC's out = 0 CCC's out = 0 |      |           |            |