

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE092619\
 Data File : BE100569.D
 Acq On : 26 Sep 2019 16:00
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 26 18:27:08 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE092019.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 20 18:26:17 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	105	0.00
2	1,4-Dioxane	0.400	0.422	-5.5	110	0.04
3	n-Nitrosodimethylamine	0.400	0.439	-9.7	152	0.02
4 S	2-Fluorophenol	0.400	0.537	-34.3#	153	0.02
5 S	Phenol-d6	0.400	0.516	-29.0#	150	0.00
6	bis(2-Chloroethyl)ether	0.400	0.410	-2.5	114	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	108	0.00
8 S	Nitrobenzene-d5	0.400	0.518	-29.5#	156	0.00
9	Naphthalene	0.400	0.387	3.3	107	0.00
10	Hexachlorobutadiene	0.400	0.379	5.3	104	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.378	5.5	108	0.00
12	2-Methylnaphthalene	0.400	0.391	2.3	112	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	97	0.00
14 S	2,4,6-Tribromophenol	0.400	0.580	-45.0#	156	0.00
15 S	2-Fluorobiphenyl	0.400	0.404	-1.0	101	0.00
16	Acenaphthylene	0.400	0.474	-18.5	127	0.01
17	Acenaphthene	0.400	0.409	-2.2	106	0.00
18	Fluorene	0.400	0.403	-0.8	105	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	87	0.00
20	4-Bromophenyl-phenylether	0.400	0.422	-5.5	95	0.00
21	Hexachlorobenzene	0.400	0.384	4.0	87	0.00
22	Pentachlorophenol	0.400	0.732	-83.0#	222	-0.01
23	Phenanthrene	0.400	0.415	-3.7	93	0.00
24	Anthracene	0.400	0.466	-16.5	112	0.00
25 SURR	Fluoranthene-d10	0.400	0.375	6.3	86	0.00
26	Fluoranthene	0.400	0.379	5.3	87	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	79	0.00
28	Pvrene	0.400	0.453	-13.2	92	0.00
29 S	Terphenyl-d14	0.400	0.421	-5.2	86	0.00
30	Benzo(a)anthracene	0.400	0.472	-18.0	98	0.01
31	Chrysene	0.400	0.391	2.3	81	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	1.200	-200.0#	262	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.443	-10.7	90	0.00
34 I	Perylene-d12	0.400	0.400	0.0	83	0.00
35	Benzo(b)fluoranthene	0.400	0.407	-1.7	93	0.00
36	Benzo(k)fluoranthene	0.400	0.354	11.5	79	0.00
37 C	Benzo(a)pyrene	0.400	0.409	-2.2	91	0.00
38	Dibenzo(a,h)anthracene	0.400	0.403	-0.8	91	0.00
39	Benzo(a,h,i)perylene	0.400	0.390	2.5	84	0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE092619\
Data File : BE100569.D
Acq On : 26 Sep 2019 16:00
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 26 18:27:08 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE092019.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 20 18:26:17 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				