

Data Path : Z:\HPCHEM1\BNA_E\Data\BE092815\
 Data File : BE090977.D
 Acq On : 28 Sep 2015 21:57
 Operator : UM/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE092815

Quant Time: Sep 30 16:02:05 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE092815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 30 15:53:58 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	204#	0.00
2	1,4-Dioxane	0.888	0.588	33.8#	158#	0.00
3	n-Nitrosodimethylamine	0.879	0.775	11.8	199#	0.00
4 S	2-Fluorophenol	1.045	0.810	22.5	180#	0.00
5 S	Phenol-d6	1.280	1.249	2.4	223#	-0.03
6	bis(2-Chloroethyl)ether	1.524	1.194	21.7	166#	-0.05
7 I	Naphthalene-d8	1.000	1.000	0.0	220#	-0.01
8 S	Nitrobenzene-d5	0.270	0.272	-0.7	234#	-0.02
9	Nitrobenzene	0.289	0.299	-3.5	254#	-0.02
10	Naphthalene	1.075	0.949	11.7	204#	0.00
11	Hexachlorobutadiene	0.179	0.150	16.2	185#	0.00
12	2-Methylnaphthalene	0.634	0.596	6.0	219#	-0.03
13 I	Acenaphthene-d10	1.000	1.000	0.0	228#	0.00
14 S	2,4,6-Tribromophenol	0.158	0.124	21.5	239#	-0.02
15 S	2-Fluorobiphenyl	1.355	1.199	11.5	217#	-0.01
16	Acenaphthylene	8.351	7.326	12.3	218#	0.00
17	Acenaphthene	1.334	1.159	13.1	211#	0.00
18	Fluorene	1.649	1.457	11.6	216#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	221#	0.00
20	4-Bromophenyl-phenylether	0.189	0.167	11.6	213#	-0.02
21	Hexachlorobenzene	0.970	0.818	15.7	195#	0.00
22	Pentachlorophenol	0.070	0.054	22.9	267#	-0.02
23	Phenanthrene	1.265	1.124	11.1	216#	-0.02
24	Anthracene	1.131	0.991	12.4	219#	-0.02
25	Fluoranthene	1.503	1.265	15.8	229#	-0.01
26 I	Chrysene-d12	1.000	1.000	0.0	221#	0.00
27	Benzydine	0.090	0.021	76.7#	62	-0.01
28	Pyrene	1.277	1.157	9.4	240#	-0.01
29 S	Terphenyl-d14	0.631	0.613	2.9	246#	-0.02
30	Benzo(a)anthracene	1.176	1.107	5.9	231#	0.00
31	3,3'-Dichlorobenzidine	0.201	0.189	6.0	284#	0.00
32	Chrysene	1.505	1.173	22.1	201#	0.00
33	Bis(2-ethylhexyl)phthalate	0.469	0.522	-11.3	386#	0.00
34	Indeno(1,2,3-cd)pyrene	1.212	1.254	-3.5	264#	-0.01
35 I	Perylene-d12	1.000	1.000	0.0	233#	0.00
36	Benzo(b)fluoranthene	1.102	1.064	3.4	251#	0.00
37	Benzo(k)fluoranthene	1.515	1.209	20.2	207#	0.00
38 C	Benzo(a)pyrene	1.022	0.963	5.8	254#	0.00
39	Dibenzo(a,h)anthracene	1.009	1.031	-2.2	283#	-0.01
40	Benzo(g,h,i)perylene	1.096	1.032	5.8	253#	-0.02

Data Path : Z:\HPCHEM1\BNA_E\Data\BE092815\
Data File : BE090977.D
Acq On : 28 Sep 2015 21:57
Operator : UM/IZ
Sample : SSTDICV001
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE092815

Quant Time: Sep 30 16:02:05 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE092815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 30 15:53:58 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				