

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE092815\
 Data File : BE090969.D
 Acq On : 28 Sep 2015 16:29
 Operator : UM/IZ
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.1

Manual Integrations
 APPROVED

MMDadoda
 9/29/2015 4:49:51 PM

Quant Time: Sep 30 16:00:07 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE092815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 29 05:34:47 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.48	152	29609	5.00	ng	0.00
7) Naphthalene-d8	10.24	136	129134	5.00	ng	0.00
13) Acenaphthene-d10	14.11	164	74507	5.00	ng	0.00
19) Phenanthrene-d10	16.86	188	176923	5.00	ng	0.00
26) Chrysene-d12	21.02	240	229273	5.00	ng	0.00
35) Perylene-d12	23.27	264	227180	5.00	ng	0.00

System Monitoring Compounds

4) 2-Fluorophenol	5.11	112	743	0.07	ng	-0.01
5) Phenol-d6	6.75	99	665	0.04	ng	0.03
8) Nitrobenzene-d5	8.67	82	481m	0.03	ng	0.02
14) 2,4,6-Tribromophenol	15.64	330	302	0.07	ng	0.00
15) 2-Fluorobiphenyl	12.75	172	2088	0.04	ng	0.00
29) Terphenyl-d14	0.00	244	0d	0.00	ng	

Target Compounds

						Qvalue
2) 1,4-Dioxane	3.07	88	652	0.06	ng	87
3) n-Nitrosodimethylamine	3.38	42	599	0.05	ng	# 79
6) bis(2-Chloroethyl)ether	6.94	93	1150m	0.07	ng	
10) Naphthalene	10.29	128	2943	0.04	ng	# 89
11) Hexachlorobutadiene	10.58	225	491	0.04	ng	95
12) 2-Methylnaphthalene	11.96	142	1574	0.04	ng	# 90
16) Acenaphthylene	13.83	152	12637	0.04	ng	99
17) Acenaphthene	14.17	154	2024	0.04	ng	99
18) Fluorene	15.19	166	2654m	0.05	ng	
20) 4-Bromophenyl-phenylether	16.07	248	808	0.05	ng	96
21) Hexachlorobenzene	16.18	284	3762	0.05	ng	98
22) Pentachlorophenol	16.58	266	390	0.09	ng	89
23) Phenanthrene	16.91	178	5813	0.06	ng	97
24) Anthracene	17.01	178	4401m	0.05	ng	
25) Fluoranthene	18.91	202	8331	0.08	ng	99
28) Pyrene	19.28	202	8669	0.07	ng	99
30) Benzo(a)anthracene	21.01	228	7137	0.06	ng	98
31) 3,3'-Dichlorobenzidine	20.99	252	898	0.05	ng	# 83
32) Chrysene	21.05	228	10007m	0.07	ng	
33) Bis(2-ethylhexyl)phthalate	20.95	149	3140	0.09	ng	# 98
34) Indeno(1,2,3-cd)pyrene	25.64	276	6448m	0.05	ng	
36) Benzo(b)fluoranthene	22.60	252	5858	0.05	ng	# 87
37) Benzo(k)fluoranthene	22.64	252	8483m	0.06	ng	
38) Benzo(a)pyrene	23.17	252	4753	0.05	ng	# 78
39) Dibenzo(a,h)anthracene	25.66	278	4751m	0.05	ng	
40) Benzo(g,h,i)perylene	26.34	276	5435	0.05	ng	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA_E\DATA\BE092815\
Data File : BE090969.D
Acq On : 28 Sep 2015 16:29
Operator : UM/IZ
Sample : SSTDIC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1

Manual Integrations
APPROVED

MMDadoda
9/29/2015 4:49:51 PM

Quant Time: Sep 30 16:00:07 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE092815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 29 05:34:47 2015
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 5.00 ng
 RT: 7.48 min Scan# 717
 Delta R.T. -0.01 min
 Lab File: BE090969.D
 Acq: 28 Sep 2015 16:29

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.1

Tgt Ion	Ratio	Lower	Upper
152	100		
150	152.3	122.6	183.8
115	64.1	52.8	79.2




Manual Integrations
 APPROVED

MMDadoda
 9/29/2015 4:49:51 PM



#2
 1,4-Dioxane
 Concen: 0.06 ng
 RT: 3.07 min Scan# 70
 Delta R.T. -0.01 min
 Lab File: BE090969.D
 Acq: 28 Sep 2015 16:29

Tgt Ion	Ratio	Lower	Upper
88	100		
58	69.6	65.8	98.6
43	37.7	25.3	37.9





#3

n-Nitrosodimethylamine

Concen: 0.05 ng

RT: 3.38 min Scan# 115

Delta R.T. -0.01 min

Lab File: BE090969.D

Acq: 28 Sep 2015 16:29

Instrument :

BNA_E

ClientSampleId :





SSTDIC0.1

Manual Integrations
APPROVED

MMDadoda

9/29/2015 4:49:51 PM

Tgt Ion:	42	Resp:	599
Ion	Ratio	Lower	Upper
42	100		
74	132.1	87.9	131.9#
44	0.0	6.6	9.8#



#4

2-Fluorophenol

Concen: 0.07 ng




RT: 5.11 min Scan# 370




Delta R.T. -0.01 min

Lab File: BE090969.D

Acq: 28 Sep 2015 16:29

Tgt Ion:	112	Resp:	743
Ion	Ratio	Lower	Upper
112	100		
64	70.5	30.9	46.3#
63	30.0	16.7	25.1#









#5
Phenol-d6
Concen: 0.04 ng
RT: 6.75 min Scan# 608
Delta R.T. 0.03 min
Lab File: BE090969.D
Acq: 28 Sep 2015 16:29

Tgt Ion: 99 Resp: 665
Ion Ratio Lower Upper
99 100
42 8.7 16.2 24.2#
71 32.8 29.0 43.4

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1


Manual Integrations
APPROVED


MMDadoda
9/29/2015 4:49:51 PM



#6
bis(2-Chloroethyl)ether
Concen: 0.07 ng m
RT: 6.94 min Scan# 637
Delta R.T. -0.03 min
Lab File: BE090969.D
Acq: 28 Sep 2015 16:29

Tgt Ion: 93 Resp: 1150
Ion Ratio Lower Upper
93 100
63 29.3 49.5 74.3#
95 6.5 22.2 33.4#





#7

Naphthalene-d8

Concen: 5.00 ng

RT: 10.24 min Scan# 1203

Delta R.T. -0.01 min

Lab File: BE090969.D


Acq: 28 Sep 2015 16:29

Instrument :

BNA_E

ClientSampleId :




SSTDICC0.1



Tgt Ion	Ratio	Resp	Lower	Upper
136	100	129134		
137	11.0	8.9	13.3	
54	10.7	8.2	12.4	
68	8.2	6.2	9.4	

Manual Integrations
APPROVED

MMDadoda
9/29/2015 4:49:51 PM



#8

Nitrobenzene-d5

Concen: 0.03 ng m




RT: 8.67 min Scan# 926


Delta R.T. 0.02 min

Lab File: BE090969.D

Acq: 28 Sep 2015 16:29

Tgt Ion	Ratio	Resp	Lower	Upper
82	100	481		
128	54.0	25.4	38.0#	
54	91.1	48.4	72.6#	









#10
Naphthalene
Concen: 0.04 ng
RT: 10.29 min Scan# 1209
Delta R.T. -0.01 min
Lab File: BE090969.D
Acq: 28 Sep 2015 16:29

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1

Tgt Ion	Ratio	Lower	Upper
128	100		
129	16.4	9.3	13.9#
127	17.6	10.7	16.1#




Manual Integrations
APPROVED


MMDadoda
9/29/2015 4:49:51 PM



#11
Hexachlorobutadiene
Concen: 0.04 ng
RT: 10.58 min Scan# 1248
Delta R.T. -0.01 min
Lab File: BE090969.D
Acq: 28 Sep 2015 16:29

Tgt Ion	Ratio	Lower	Upper
225	100		
223	67.2	49.0	73.6
227	65.4	50.3	75.5





#12

2-Methylnaphthalene

Concen: 0.04 ng

RT: 11.96 min Scan# 1484

Delta R.T. 0.01 min

Lab File: BE090969.D

Acq: 28 Sep 2015 16:29

Instrument :

BNA_E





ClientSampleId :

SSTDICC0.1

Tgt Ion:	142	Resp:	1574
Ion Ratio	Lower	Upper	
142	100		
141	85.2	71.4	107.0
115	51.7	30.3	45.5

Manual Integrations
APPROVED

MMDadoda
9/29/2015 4:49:51 PM



#13

Acenaphthene-d10

Concen: 5.00 ng




RT: 14.11 min Scan# 1906


Delta R.T. -0.00 min

Lab File: BE090969.D

Acq: 28 Sep 2015 16:29

Tgt Ion:	164	Resp:	74507
Ion Ratio	Lower	Upper	
164	100		
162	103.6	85.3	127.9
160	51.7	46.2	69.2









#14
 2,4,6-Tribromophenol
 Concen: 0.07 ng
 RT: 15.64 min Scan# 2047
 Delta R.T. 0.00 min
 Lab File: BE090969.D
 Acq: 28 Sep 2015 16:29

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.1

Tgt Ion	Ratio	Lower	Upper
330	100		
332	94.7	79.1	118.7
141	253.6	125.4	188.0#




Manual Integrations
 APPROVED


MMDadoda
 9/29/2015 4:49:51 PM



#15
 2-Fluorobiphenyl
 Concen: 0.04 ng
 RT: 12.75 min Scan# 1657
 Delta R.T. 0.00 min
 Lab File: BE090969.D
 Acq: 28 Sep 2015 16:29

Tgt Ion	Ratio	Lower	Upper
172	100		
171	38.6	28.8	43.2
170	29.6	19.3	28.9#





#16

Acenaphthylene

Concen: 0.04 ng

RT: 13.83 min Scan# 1863

Delta R.T. -0.00 min

Lab File: BE090969.D

Acq: 28 Sep 2015 16:29

Instrument :

BNA_E





ClientSampleId :

SSTDIC0.1

Tgt Ion:	152	Resp:	12637
Ion Ratio	Lower	Upper	
152	100		
151	4.7	3.9	5.9
153	13.2	10.3	15.5

Manual Integrations
APPROVED

MMDadoda
9/29/2015 4:49:51 PM



#17

Acenaphthene

Concen: 0.04 ng




RT: 14.17 min Scan# 1916


Delta R.T. -0.00 min

Lab File: BE090969.D

Acq: 28 Sep 2015 16:29

Tgt Ion:	154	Resp:	2024
Ion Ratio	Lower	Upper	
154	100		
153	439.3	352.0	528.0
152	256.3	200.0	300.0









#18
Fluorene
 Concen: 0.05 ng m
 RT: 15.19 min Scan# 2021
 Delta R.T. 0.02 min
 Lab File: BE090969.D
 Acq: 28 Sep 2015 16:29

Instrument :
 BNA_E
 ClientSampled :
 SSTDICC0.1

Tgt Ion	Ratio	Lower	Upper
166	100		
165	109.2	80.8	121.2
167	14.5	10.9	16.3




Manual Integrations
 APPROVED


MMDadoda
 9/29/2015 4:49:51 PM



#19
Phenanthrene-d10
 Concen: 5.00 ng
 RT: 16.86 min Scan# 2117
 Delta R.T. 0.00 min
 Lab File: BE090969.D
 Acq: 28 Sep 2015 16:29

Tgt Ion	Ratio	Lower	Upper
188	100		
94	11.0	8.7	13.1
80	12.0	9.4	14.0









#20
4-Bromophenyl-phenylether
Concen: 0.05 ng
RT: 16.07 min Scan# 2072
Delta R.T. 0.00 min
Lab File: BE090969.D
Acq: 28 Sep 2015 16:29

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1

Tgt Ion	Ratio	Lower	Upper
248	100		
250	97.6	80.5	120.7
141	335.6	261.9	392.9




Manual Integrations
APPROVED


MMDadoda
9/29/2015 4:49:51 PM



#21
Hexachlorobenzene
Concen: 0.05 ng
RT: 16.18 min Scan# 2078
Delta R.T. 0.00 min
Lab File: BE090969.D
Acq: 28 Sep 2015 16:29

Tgt Ion	Ratio	Lower	Upper
284	100		
142	41.8	32.2	48.2
249	31.7	25.4	38.2





#22

Pentachlorophenol

Concen: 0.09 ng

RT: 16.58 min Scan# 2101

Delta R.T. 0.02 min

Lab File: BE090969.D

Acq: 28 Sep 2015 16:29

Instrument :

BNA_E





ClientSampleId :

SSTDICC0.1

Tgt Ion:	266	Resp:	390
Ion Ratio	Lower	Upper	
266	100		
268	82.2	58.2	87.2
264	79.1	56.2	84.4

Manual Integrations
APPROVED

MMDadoda
9/29/2015 4:49:51 PM



#23

Phenanthrene

Concen: 0.06 ng




RT: 16.91 min Scan# 2120


Delta R.T. 0.00 min

Lab File: BE090969.D

Acq: 28 Sep 2015 16:29

Tgt Ion:	178	Resp:	5813
Ion Ratio	Lower	Upper	
178	100		
176	19.1	15.3	22.9
179	18.7	12.6	18.8









#24
 Anthracene
 Concen: 0.05 ng m
 RT: 17.01 min Scan# 2126
 Delta R.T. 0.00 min
 Lab File: BE090969.D
 Acq: 28 Sep 2015 16:29

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.1

Tgt Ion	Ratio	Lower	Upper
178	100		
176	14.9	14.6	21.8
179	10.9	11.8	17.6




Manual Integrations
 APPROVED


MMDadoda
 9/29/2015 4:49:51 PM



#25
 Fluoranthene
 Concen: 0.08 ng
 RT: 18.91 min Scan# 2293
 Delta R.T. 0.00 min
 Lab File: BE090969.D
 Acq: 28 Sep 2015 16:29

Tgt Ion	Ratio	Lower	Upper
202	100		
101	13.1	11.0	16.6
203	17.3	13.9	20.9









#26
Chrysene-d12
Concen: 5.00 ng
RT: 21.02 min Scan# 2551
Delta R.T. -0.00 min
Lab File: BE090969.D
Acq: 28 Sep 2015 16:29

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1

Tgt Ion	Ratio	Resp	Lower	Upper
240	100	229273		
120	16.5	11.0	16.4#	
236	28.1	21.7	32.5	




Manual Integrations
APPROVED


MMDadoda
9/29/2015 4:49:51 PM



#28
Pyrene
Concen: 0.07 ng
RT: 19.28 min Scan# 2341
Delta R.T. 0.00 min
Lab File: BE090969.D
Acq: 28 Sep 2015 16:29

Tgt Ion	Ratio	Resp	Lower	Upper
202	100	8669		
200	20.7	16.8	25.2	
203	17.8	14.0	21.0	





#29

Terphenyl-d14

Concen: 0.00 ng

Expected RT: 19.50 min

Lab File: BE090969.D

Acq: 28 Sep 2015 16:29

Tgt Ion: 244

Sig Exp Ratio

244 100

212 8.6

122 13.7

Instrument :



BNA_E

ClientSampleId :

SSTDICC0.1

Manual Integrations
APPROVED

MMDadoda
9/29/2015 4:49:51 PM



#30

Benzo(a)anthracene

Concen: 0.06 ng

RT: 21.01 min Scan# 2550

Delta R.T. -0.00 min

Lab File: BE090969.D

Acq: 28 Sep 2015 16:29




Tgt Ion: 228 Resp: 7137


Ion Ratio Lower Upper

228 100

226 27.0 21.0 31.4

229 21.2 15.7 23.5









#31
 3,3'-Dichlorobenzidine
 Concen: 0.05 ng
 RT: 20.99 min Scan# 2548
 Delta R.T. 0.02 min
 Lab File: BE090969.D
 Acq: 28 Sep 2015 16:29

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.1

Tgt Ion	Ratio	Lower	Upper
252	100		
254	66.5	51.1	76.7
126	45.6	12.6	18.8




Manual Integrations
 APPROVED


MMDadoda
 9/29/2015 4:49:51 PM



#32
 Chrysene
 Concen: 0.07 ng m
 RT: 21.05 min Scan# 2555
 Delta R.T. -0.00 min
 Lab File: BE090969.D
 Acq: 28 Sep 2015 16:29

Tgt Ion	Ratio	Lower	Upper
228	100		
226	31.3	24.0	36.0
229	19.6	15.9	23.9









#33
 Bis(2-ethylhexyl)phthalate
 Concen: 0.09 ng
 RT: 20.95 min Scan# 2544
 Delta R.T. -0.00 min
 Lab File: BE090969.D
 Acq: 28 Sep 2015 16:29

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.1

Tgt Ion	Ratio	Lower	Upper
149	100		
167	25.4	19.5	29.3
279	3.7	2.3	3.5




Manual Integrations
 APPROVED


MMDadoda
 9/29/2015 4:49:51 PM



#34
 Indeno(1,2,3-cd)pyrene
 Concen: 0.05 ng m
 RT: 25.64 min Scan# 3463
 Delta R.T. 0.04 min
 Lab File: BE090969.D
 Acq: 28 Sep 2015 16:29

Tgt Ion	Ratio	Lower	Upper
276	100		
138	27.2	23.4	35.2
277	21.1	19.8	29.8









#35
 Perylene-d12
 Concen: 5.00 ng
 RT: 23.27 min Scan# 2946
 Delta R.T. -0.00 min
 Lab File: BE090969.D
 Acq: 28 Sep 2015 16:29

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC0.1

Tgt Ion	Ratio	Resp	Lower	Upper
264	100	227180		
260	25.1	20.0	30.0	
265	21.7	17.5	26.3	




Manual Integrations
 APPROVED


MMDadoda
 9/29/2015 4:49:51 PM



#36
 Benzo(b)fluoranthene
 Concen: 0.05 ng
 RT: 22.60 min Scan# 2798
 Delta R.T. 0.00 min
 Lab File: BE090969.D
 Acq: 28 Sep 2015 16:29

Tgt Ion	Ratio	Resp	Lower	Upper
252	100	5858		
253	26.8	17.4	26.0#	
125	19.5	10.2	15.2#	









#37
Benzo(k)fluoranthene
Concen: 0.06 ng m
RT: 22.64 min Scan# 2808
Delta R.T. 0.00 min
Lab File: BE090969.D
Acq: 28 Sep 2015 16:29

Instrument :
BNA_E
ClientSampleId :
SSTDIC0.1

Tgt Ion	Ratio	Lower	Upper
252	100		
253	26.6	17.6	26.4#
125	15.9	9.9	14.9#




Manual Integrations
APPROVED


MMDadoda
9/29/2015 4:49:51 PM



#38
Benzo(a)pyrene
Concen: 0.05 ng
RT: 23.17 min Scan# 2924
Delta R.T. -0.00 min
Lab File: BE090969.D
Acq: 28 Sep 2015 16:29

Tgt Ion	Ratio	Lower	Upper
252	100		
253	31.1	18.1	27.1#
125	25.2	11.0	16.4#





#39

Dibenzo(a,h)anthracene

Concen: 0.05 ng m

RT: 25.66 min Scan# 3467

Delta R.T. 0.04 min

Lab File: BE090969.D


Acq: 28 Sep 2015 16:29

Instrument :

BNA_E

ClientSampleId :

SSTDIC0.1



Tgt Ion: 278 Resp: 4751

Ion Ratio Lower Upper




278 100

139 32.2 16.0 24.0#

279 30.8 19.4 29.2#

Manual Integrations
APPROVED

MMDadoda
9/29/2015 4:49:51 PM



#40

Benzo(g,h,i)perylene

Concen: 0.05 ng

RT: 26.34 min Scan# 3615

Delta R.T. 0.02 min

Lab File: BE090969.D

Acq: 28 Sep 2015 16:29

Tgt Ion: 276 Resp: 5435

Ion Ratio Lower Upper

276 100

277 27.2 18.6 28.0

138 30.5 20.4 30.6

