

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE092815\
 Data File : BE090977.D
 Acq On : 28 Sep 2015 21:57
 Operator : UM/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE092815

Manual Integrations
 APPROVED

MMDadoda
 9/29/2015 4:50:55 PM

Quant Time: Sep 30 16:02:05 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE092815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 30 15:53:58 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.48	152	70213	5.00	ng	0.00
7) Naphthalene-d8	10.24	136	344285	5.00	ng	-0.01
13) Acenaphthene-d10	14.11	164	197650	5.00	ng	0.00
19) Phenanthrene-d10	16.86	188	450811	5.00	ng	0.00
26) Chrysene-d12	21.02	240	545315	5.00	ng	0.00
35) Perylene-d12	23.27	264	516946	5.00	ng	0.00

System Monitoring Compounds						
4) 2-Fluorophenol	5.11	112	11368	0.86	ng	0.00
5) Phenol-d6	6.68	99	17537	0.98	ng	-0.03
8) Nitrobenzene-d5	8.64	82	18752	0.98	ng	-0.02
14) 2,4,6-Tribromophenol	15.62	330	4909	0.87	ng	-0.02
15) 2-Fluorobiphenyl	12.73	172	47386	0.88	ng	-0.01
29) Terphenyl-d14	19.48	244	66897	1.02	ng	-0.02

Target Compounds				Qvalue		
2) 1,4-Dioxane	3.09	88	8262	0.71	ng	98
3) n-Nitrosodimethylamine	3.38	42	10876	0.96	ng	93
6) bis(2-Chloroethyl)ether	6.92	93	16762	0.91	ng	# 75
9) Nitrobenzene	8.68	77	20578	1.09	ng	98
10) Naphthalene	10.29	128	65356	0.92	ng	99
11) Hexachlorobutadiene	10.58	225	10299	0.85	ng	98
12) 2-Methylnaphthalene	11.92	142	41006	0.94	ng	99
16) Acenaphthylene	13.82	152	289594	0.88	ng	100
17) Acenaphthene	14.17	154	45823	0.87	ng	98
18) Fluorene	15.17	166	57597	0.88	ng	100
20) 4-Bromophenyl-phenylether	16.06	248	15081	0.89	ng	92
21) Hexachlorobenzene	16.18	284	73735	0.87	ng	100
22) Pentachlorophenol	16.54	266	4878	0.99	ng	92
23) Phenanthrene	16.89	178	101358	0.93	ng	100
24) Anthracene	17.00	178	89322	0.92	ng	99
25) Fluoranthene	18.90	202	114012	0.94	ng	100
27) Benzidine	19.13	184	2238	0.20	ng	93
28) Pyrene	19.26	202	126136	1.00	ng	100
30) Benzo(a)anthracene	21.00	228	120697	1.00	ng	98
31) 3,3'-Dichlorobenzidine	20.96	252	20591	1.11	ng	# 96
32) Chrysene	21.05	228	127929	0.87	ng	100
33) Bis(2-ethylhexyl)phthalate	20.95	149	56956	1.37	ng	98
34) Indeno(1,2,3-cd)pyrene	25.59	276	136755	1.09	ng	99
36) Benzo(b)fluoranthene	22.59	252	110039	0.97	ng	99
37) Benzo(k)fluoranthene	22.63	252	125035	0.83	ng	100
38) Benzo(a)pyrene	23.17	252	99574	0.94	ng	98
39) Dibenzo(a,h)anthracene	25.61	278	106571	1.02	ng	99
40) Benzo(g,h,i)perylene	26.30	276	106731	0.94	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA_E\DATA\BE092815\
Data File : BE090977.D
Acq On : 28 Sep 2015 21:57
Operator : UM/IZ
Sample : SSTDICV001
Misc :
ALS Vial : 10 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
ICVBE092815

Manual Integrations
APPROVED

MMDadoda
9/29/2015 4:50:55 PM

Quant Time: Sep 30 16:02:05 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE092815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 30 15:53:58 2015
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 7.48 min Scan# 717

Delta R.T. -0.01 min

Lab File: BE090977.D


Acq: 28 Sep 2015 21:57

Instrument :

BNA_E

ClientSampleId :




ICVBE092815



Tgt Ion:	152	Resp:	70213
Ion Ratio	Lower	Upper	
152	100		
150	154.1	122.6	183.8
115	65.4	52.8	79.2

Manual Integrations
APPROVED

MMDadoda
9/29/2015 4:50:55 PM



#2

1,4-Dioxane

Concen: 0.71 ng




RT: 3.09 min Scan# 72


Delta R.T. -0.00 min

Lab File: BE090977.D

Acq: 28 Sep 2015 21:57

Tgt Ion:	88	Resp:	8262
Ion Ratio	Lower	Upper	
88	100		
58	83.4	65.8	98.6
43	30.1	25.3	37.9





#3





n-Nitrosodimethylamine
 Concen: 0.96 ng
 RT: 3.38 min Scan# 116
 Delta R.T. -0.00 min
 Lab File: BE090977.D
 Acq: 28 Sep 2015 21:57

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE092815

Manual Integrations
 APPROVED

MMDadoda
 9/29/2015 4:50:55 PM




Tgt Ion:	42	Resp:	10876
Ion	Ratio	Lower	Upper
42	100		
74	117.3	87.9	131.9
44	7.8	6.6	9.8






#4

2-Fluorophenol
 Concen: 0.86 ng
 RT: 5.11 min Scan# 371
 Delta R.T. -0.01 min
 Lab File: BE090977.D
 Acq: 28 Sep 2015 21:57

Tgt Ion:	112	Resp:	11368
Ion	Ratio	Lower	Upper
112	100		
64	71.0	30.9	46.3#
63	36.0	16.7	25.1#





#5

Phenol-d6

Concen: 0.98 ng

RT: 6.68 min Scan# 599

Delta R.T. -0.03 min

Lab File: BE090977.D

Acq: 28 Sep 2015 21:57

Tgt Ion:	99	Resp:	17537
Ion	Ratio	Lower	Upper
99	100		
42	21.2	16.2	24.2
71	35.4	29.0	43.4

Instrument :





BNA_E

ClientSampleId :

ICVBE092815

Manual Integrations
APPROVED

MMDadoda
9/29/2015 4:50:55 PM



#6

bis(2-Chloroethyl)ether

Concen: 0.91 ng


RT: 6.92 min Scan# 634


Delta R.T. -0.05 min

Lab File: BE090977.D

Acq: 28 Sep 2015 21:57

Tgt Ion:	93	Resp:	16762
Ion	Ratio	Lower	Upper
93	100		
63	85.5	49.5	74.3#
95	33.9	22.2	33.4#





#7

Naphthalene-d8

Concen: 5.00 ng

RT: 10.24 min Scan# 1203

Delta R.T. -0.01 min

Lab File: BE090977.D


Acq: 28 Sep 2015 21:57

Instrument :

BNA_E

ClientSampleId :




ICVBE092815



Tgt Ion	Ratio	Resp	Lower	Upper
136	100	344285		
137	11.1	8.9	13.3	
54	11.1	8.2	12.4	
68	8.2	6.2	9.4	

Manual Integrations
APPROVED

MMDadoda
9/29/2015 4:50:55 PM



#8

Nitrobenzene-d5

Concen: 0.98 ng




RT: 8.64 min Scan# 918


Delta R.T. -0.02 min

Lab File: BE090977.D

Acq: 28 Sep 2015 21:57

Tgt Ion	Ratio	Resp	Lower	Upper
82	100	18752		
128	29.8	25.4	38.0	
54	55.9	48.4	72.6	





#9

Nitrobenzene

Concen: 1.09 ng

RT: 8.68 min Scan# 927

Delta R.T. -0.02 min

Lab File: BE090977.D


Acq: 28 Sep 2015 21:57

Instrument :

BNA_E

ClientSampleId :




ICVBE092815



Tgt Ion:	77	Resp:	20578
Ion	Ratio	Lower	Upper
77	100		
123	32.3	26.9	40.3
65	12.0	9.4	14.2

Manual Integrations
APPROVED

MMDadoda
9/29/2015 4:50:55 PM



#10

Naphthalene

Concen: 0.92 ng




RT: 10.29 min Scan# 1210


Delta R.T. -0.01 min

Lab File: BE090977.D

Acq: 28 Sep 2015 21:57


Tgt Ion:	128	Resp:	65356
Ion	Ratio	Lower	Upper
128	100		
129	11.3	9.3	13.9
127	12.9	10.7	16.1





#11
Hexachlorobutadiene
Concen: 0.85 ng
RT: 10.58 min Scan# 1249
Delta R.T. -0.01 min
Lab File: BE090977.D
Acq: 28 Sep 2015 21:57




Instrument :
BNA_E
ClientSampleId :
ICVBE092815




Tgt Ion	Ratio	Resp	Lower	Upper
225	100	10299		
223	63.6	49.0	73.6	
227	64.4	50.3	75.5	

Manual Integrations
APPROVED



MMDadoda
9/29/2015 4:50:55 PM




#12
2-Methylnaphthalene
Concen: 0.94 ng
RT: 11.92 min Scan# 1476
Delta R.T. -0.03 min
Lab File: BE090977.D
Acq: 28 Sep 2015 21:57



Tgt Ion	Ratio	Resp	Lower	Upper
142	100	41006		
141	88.1	71.4	107.0	
115	37.2	30.3	45.5	









#13
Acenaphthene-d10
Concen: 5.00 ng
RT: 14.11 min Scan# 1907
Delta R.T. -0.00 min
Lab File: BE090977.D
Acq: 28 Sep 2015 21:57

Instrument :
BNA_E
ClientSampleId :
ICVBE092815

Tgt Ion	Ratio	Lower	Upper
164	100		
162	104.3	85.3	127.9
160	51.0	46.2	69.2




Manual Integrations
APPROVED


MMDadoda
9/29/2015 4:50:55 PM



#14
2,4,6-Tribromophenol
Concen: 0.87 ng
RT: 15.62 min Scan# 2048
Delta R.T. -0.02 min
Lab File: BE090977.D
Acq: 28 Sep 2015 21:57

Tgt Ion	Ratio	Lower	Upper
330	100		
332	96.0	79.1	118.7
141	160.2	125.4	188.0









#15
2-Fluorobiphenyl
Concen: 0.88 ng
RT: 12.73 min Scan# 1655
Delta R.T. -0.01 min
Lab File: BE090977.D
Acq: 28 Sep 2015 21:57

Instrument :
BNA_E
ClientSampleId :
ICVBE092815

Tgt Ion	Ratio	Lower	Upper
172	100		
171	34.3	28.8	43.2
170	23.4	19.3	28.9




Manual Integrations
APPROVED


MMDadoda
9/29/2015 4:50:55 PM



#16
Acenaphthylene
Concen: 0.88 ng
RT: 13.82 min Scan# 1863
Delta R.T. -0.01 min
Lab File: BE090977.D
Acq: 28 Sep 2015 21:57

Tgt Ion	Ratio	Lower	Upper
152	100		
151	4.9	3.9	5.9
153	12.9	10.3	15.5









#17
Acenaphthene
Concen: 0.87 ng
RT: 14.17 min Scan# 1916
Delta R.T. -0.01 min
Lab File: BE090977.D
Acq: 28 Sep 2015 21:57

Instrument :
BNA_E
ClientSampleId :
ICVBE092815

Tgt Ion	Ratio	Lower	Upper
154	100		
153	441.6	352.0	528.0
152	240.8	200.0	300.0




Manual Integrations
APPROVED


MMDadoda
9/29/2015 4:50:55 PM



#18
Fluorene
Concen: 0.88 ng
RT: 15.17 min Scan# 2022
Delta R.T. 0.00 min
Lab File: BE090977.D
Acq: 28 Sep 2015 21:57

Tgt Ion	Ratio	Lower	Upper
166	100		
165	100.5	80.8	121.2
167	13.7	10.9	16.3









#19
Phenanthrene-d10
Concen: 5.00 ng
RT: 16.86 min Scan# 2119
Delta R.T. 0.00 min
Lab File: BE090977.D
Acq: 28 Sep 2015 21:57

Instrument :
BNA_E
ClientSampleId :
ICVBE092815

Tgt Ion	Ratio	Resp Lower	Upper
188	100		
94	6.9	8.7	13.1#
80	7.2	9.4	14.0#




Manual Integrations
APPROVED


MMDadoda
9/29/2015 4:50:55 PM



#20
4-Bromophenyl-phenylether
Concen: 0.89 ng
RT: 16.06 min Scan# 2073
Delta R.T. -0.02 min
Lab File: BE090977.D
Acq: 28 Sep 2015 21:57

Tgt Ion	Ratio	Resp Lower	Upper
248	100		
250	95.8	80.5	120.7
141	345.1	261.9	392.9









#21
Hexachlorobenzene
Concen: 0.87 ng
RT: 16.18 min Scan# 2080
Delta R.T. 0.00 min
Lab File: BE090977.D
Acq: 28 Sep 2015 21:57

Instrument :
BNA_E
ClientSampleId :
ICVBE092815

Tgt Ion	Ratio	Lower	Upper
284	100		
142	40.0	32.2	48.2
249	31.6	25.4	38.2




Manual Integrations
APPROVED


MMDadoda
9/29/2015 4:50:55 PM



#22
Pentachlorophenol
Concen: 0.99 ng
RT: 16.54 min Scan# 2101
Delta R.T. -0.02 min
Lab File: BE090977.D
Acq: 28 Sep 2015 21:57

Tgt Ion	Ratio	Lower	Upper
266	100		
268	68.1	58.2	87.2
264	61.3	56.2	84.4









#23
Phenanthrene
Concen: 0.93 ng
RT: 16.89 min Scan# 2121
Delta R.T. -0.02 min
Lab File: BE090977.D
Acq: 28 Sep 2015 21:57

Instrument :
BNA_E
ClientSampleId :
ICVBE092815

Tgt Ion	Ratio	Lower	Upper
178	100		
176	19.0	15.3	22.9
179	15.8	12.6	18.8




Manual Integrations
APPROVED


MMDadoda
9/29/2015 4:50:55 PM



#24
Anthracene
Concen: 0.92 ng
RT: 17.00 min Scan# 2127
Delta R.T. -0.02 min
Lab File: BE090977.D
Acq: 28 Sep 2015 21:57

Tgt Ion	Ratio	Lower	Upper
178	100		
176	18.6	14.6	21.8
179	15.2	11.8	17.6





#25

Fluoranthene

Concen: 0.94 ng

RT: 18.90 min Scan# 2293

Delta R.T. -0.01 min

Lab File: BE090977.D

Acq: 28 Sep 2015 21:57

Instrument :

BNA_E


ClientSampleId :

ICVBE092815




Tgt	Ion	202	Resp:	114012
Ion	Ratio	Lower	Upper	
202	100			
101	13.8	11.0	16.6	
203	17.3	13.9	20.9	

Manual Integrations
APPROVED

MMDadoda
9/29/2015 4:50:55 PM



Abundance Ion 202.00 (201.70 to 202.70): E
Ion 101.00 (100.70 to 101.70): E
Ion 203.00 (202.70 to 203.70): E



#26

Chrysene-d12

Concen: 5.00 ng


RT: 21.02 min Scan# 2553

Delta R.T. -0.00 min



Lab File: BE090977.D


Acq: 28 Sep 2015 21:57

Tgt	Ion	240	Resp:	545315
Ion	Ratio	Lower	Upper	
240	100			
120	12.5	11.0	16.4	
236	27.0	21.7	32.5	



Abundance Ion 240.00 (239.70 to 240.70): E
Ion 120.00 (119.70 to 120.70): E
Ion 236.00 (235.70 to 236.70): E





#27

Benzidine

Concen: 0.20 ng

RT: 19.13 min Scan# 2324

Delta R.T. -0.01 min

Lab File: BE090977.D

Acq: 28 Sep 2015 21:57

Instrument :

BNA_E





ClientSampleId :

ICVBE092815

Manual Integrations
APPROVED

MMDadoda
9/29/2015 4:50:55 PM

Tgt Ion:	184	Resp:	2238
Ion	Ratio	Lower	Upper
184	100		
185	24.5	22.3	33.5
183	17.3	16.4	24.6



#28

Pyrene

Concen: 1.00 ng




RT: 19.26 min Scan# 2341


Delta R.T. -0.01 min

Lab File: BE090977.D

Acq: 28 Sep 2015 21:57

Tgt Ion:	202	Resp:	126136
Ion	Ratio	Lower	Upper
202	100		
200	21.0	16.8	25.2
203	17.5	14.0	21.0









#29
 Terphenyl-d14
 Concen: 1.02 ng
 RT: 19.48 min Scan# 2370
 Delta R.T. -0.02 min
 Lab File: BE090977.D
 Acq: 28 Sep 2015 21:57

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE092815

Tgt Ion	Ratio	Resp Lower	Resp Upper
244	100		
212	8.2	6.9	10.3
122	14.2	11.0	16.4




Manual Integrations
 APPROVED


MMDadoda
 9/29/2015 4:50:55 PM



#30
 Benzo(a)anthracene
 Concen: 1.00 ng
 RT: 21.00 min Scan# 2551
 Delta R.T. -0.01 min
 Lab File: BE090977.D
 Acq: 28 Sep 2015 21:57

Tgt Ion	Ratio	Resp Lower	Resp Upper
228	100		
226	27.5	21.0	31.4
229	19.5	15.7	23.5









#31
 3,3'-Dichlorobenzidine
 Concen: 1.11 ng
 RT: 20.96 min Scan# 2547
 Delta R.T. -0.01 min
 Lab File: BE090977.D
 Acq: 28 Sep 2015 21:57

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE092815

Tgt Ion	Ratio	Lower	Upper
252	100		
254	65.9	51.1	76.7
126	11.4	12.6	18.8#




Manual Integrations
 APPROVED


MMDadoda
 9/29/2015 4:50:55 PM



#32
 Chrysene
 Concen: 0.87 ng
 RT: 21.05 min Scan# 2557
 Delta R.T. -0.00 min
 Lab File: BE090977.D
 Acq: 28 Sep 2015 21:57


Tgt Ion	Ratio	Lower	Upper
228	100		
226	29.7	24.0	36.0
229	19.9	15.9	23.9





#33
 Bis(2-ethylhexyl)phthalate
 Concen: 1.37 ng
 RT: 20.95 min Scan# 2546
 Delta R.T. -0.00 min
 Lab File: BE090977.D
 Acq: 28 Sep 2015 21:57




Instrument :
 BNA_E
 ClientSampleId :
 ICVBE092815



Tgt Ion	Ratio	Resp Lower	Resp Upper
149	100		
167	25.7	19.5	29.3
279	2.8	2.3	3.5




Manual Integrations
 APPROVED


MMDadoda
 9/29/2015 4:50:55 PM



#34
 Indeno(1,2,3-cd)pyrene
 Concen: 1.09 ng
 RT: 25.59 min Scan# 3454
 Delta R.T. -0.01 min
 Lab File: BE090977.D
 Acq: 28 Sep 2015 21:57


Tgt Ion	Ratio	Resp Lower	Resp Upper
276	100		
138	29.9	23.4	35.2
277	24.8	19.8	29.8





#35
 Perylene-d12
 Concen: 5.00 ng
 RT: 23.27 min Scan# 2948
 Delta R.T. -0.00 min
 Lab File: BE090977.D
 Acq: 28 Sep 2015 21:57




Instrument :
 BNA_E
 ClientSampleId :
 ICVBE092815



Tgt Ion: 264 Resp: 516946
 Ion Ratio Lower Upper
 264 100
 260 25.1 20.0 30.0
 265 21.8 17.5 26.3




Manual Integrations
 APPROVED


MMDadoda
 9/29/2015 4:50:55 PM



#36
 Benzo(b)fluoranthene
 Concen: 0.97 ng
 RT: 22.59 min Scan# 2798
 Delta R.T. -0.00 min
 Lab File: BE090977.D
 Acq: 28 Sep 2015 21:57

Tgt Ion: 252 Resp: 110039
 Ion Ratio Lower Upper
 252 100
 253 21.9 17.4 26.0
 125 13.1 10.2 15.2





#37

Benzo(k)fluoranthene

Concen: 0.83 ng

RT: 22.63 min Scan# 2808

Delta R.T. -0.00 min

Lab File: BE090977.D


Acq: 28 Sep 2015 21:57

Instrument :

BNA_E

ClientSampleId :




ICVBE092815



Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.2	17.6	26.4
125	12.5	9.9	14.9

Manual Integrations
APPROVED

MMDadoda
9/29/2015 4:50:55 PM



#38

Benzo(a)pyrene


Concen: 0.94 ng

RT: 23.17 min Scan# 2925



Delta R.T. -0.00 min


Lab File: BE090977.D

Acq: 28 Sep 2015 21:57



Tgt Ion	Ratio	Lower	Upper
252	100		
253	21.8	18.1	27.1
125	14.4	11.0	16.4





#39

Dibenzo(a,h)anthracene

Concen: 1.02 ng

RT: 25.61 min Scan# 3459

Delta R.T. -0.01 min

Lab File: BE090977.D


Acq: 28 Sep 2015 21:57

Instrument :

BNA_E

ClientSampleId :



ICVBE092815



Tgt Ion	Ratio	Resp	Lower	Upper
278	100	106571		
139	18.7	16.0	24.0	
279	24.2	19.4	29.2	

Manual Integrations
APPROVED

MMDadoda
9/29/2015 4:50:55 PM



#40

Benzo(g,h,i)perylene


Concen: 0.94 ng

RT: 26.30 min Scan# 3609

Delta R.T. -0.02 min

Lab File: BE090977.D

Acq: 28 Sep 2015 21:57



Tgt Ion	Ratio	Resp	Lower	Upper
276	100	106731		
277	23.6	18.6	28.0	
138	25.5	20.4	30.6	

