

Data Path : Z:\HPCHEM1\BNA E\Data\BE092817\
 Data File : BE094134.D
 Acq On : 28 Sep 2017 17:36
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 28 18:10:53 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE092517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Sep 25 17:37:08 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	147	0.00
2	1,4-Dioxane	0.780	0.749	4.0	136	0.01
3	n-Nitrosodimethylamine	0.996	0.785	21.2	116	0.01
4 S	2-Fluorophenol	1.516	1.389	8.4	133	0.00
5 S	Phenol-d6	2.106	2.003	4.9	138	0.00
6	bis(2-Chloroethyl)ether	1.632	1.574	3.6	139	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	152#	0.00
8 S	Nitrobenzene-d5	0.387	0.330	14.7	137	0.00
9	Naphthalene	4.516	4.521	-0.1	153#	0.00
10	Hexachlorobutadiene	0.169	0.171	-1.2	160#	0.00
11 SURR	2-Methylnaphthalene-d10	0.616	0.617	-0.2	155#	0.00
12	2-Methylnaphthalene	0.731	0.730	0.1	156#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	153#	0.00
14 S	2,4,6-Tribromophenol	0.256	0.161	37.1#	112	0.00
15 S	2-Fluorobiphenyl	1.322	1.384	-4.7	153#	0.00
16	Acenaphthylene	2.005	2.015	-0.5	149	0.00
17	Acenaphthene	1.253	1.317	-5.1	150	0.00
18	Fluorene	1.700	1.719	-1.1	150	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	131	0.00
20	4-Bromophenyl-phenylether	0.164	0.175	-6.7	144	0.00
21	Hexachlorobenzene	0.104	0.129	-24.0	166#	0.00
22	Pentachlorophenol	0.095	0.058	38.9#	83	0.00
23	Phenanthrene	1.023	1.100	-7.5	147	0.00
24	Anthracene	1.074	1.074	0.0	139	0.00
25 SURR	Fluoranthene-d10	3.873	3.542	8.5	126	0.00
26	Fluoranthene	1.147	1.057	7.8	125	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	111	0.00
28	Pvrene	1.267	1.341	-5.8	123	0.00
29 S	Terphenyl-d14	0.730	0.739	-1.2	119	0.00
30	Benzo(a)anthracene	1.347	1.302	3.3	108	0.00
31	Chrysene	1.300	1.292	0.6	111	0.00
32	Bis(2-ethylhexyl)phthalate	6.447	3.354	48.0#	78	0.00
33	Indeno(1,2,3-cd)pyrene	1.391	1.281	7.9	107	0.00
34 I	Perylene-d12	1.000	1.000	0.0	109	0.00
35	Benzo(b)fluoranthene	1.443	1.297	10.1	108	0.00
36	Benzo(k)fluoranthene	1.382	1.338	3.2	105	0.00
37 C	Benzo(a)pyrene	1.304	1.249	4.2	108	0.00
38	Dibenzo(a,h)anthracene	1.235	1.143	7.4	106	0.00
39	Benzo(a,h,i)perylene	1.222	1.150	5.9	106	0.00

Data Path : Z:\HPCHEM1\BNA E\Data\BE092817\
Data File : BE094134.D
Acq On : 28 Sep 2017 17:36
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Sep 28 18:10:53 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE092517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Sep 25 17:37:08 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			