


Data Path : Z:\HPCHEM1\BNA E\Data\BE092817\  
 Data File : BE094133.D  
 Acq On : 28 Sep 2017 16:58  
 Operator : SJ/JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_E  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_E\METHODS\8270-SIM-BE092517.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Fri Sep 29 13:57:50 2017



AutoFind: Scans 2567, 2568, 2569; Background Corrected with Scan 2557

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 51.4      | 14727   | PASS             |
| 68          | 69           | 0.00         | 2            | 0.3       | 40      | PASS             |
| 69          | 198          | 0.00         | 100          | 47.6      | 13641   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.6       | 78      | PASS             |
| 127         | 198          | 10           | 80           | 51.8      | 14856   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.0       | 0       | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 28675   | PASS             |
| 199         | 198          | 5            | 9            | 7.0       | 2015    | PASS             |
| 275         | 198          | 10           | 60           | 22.5      | 6466    | PASS             |
| 365         | 198          | 1            | 100          | 2.4       | 696     | PASS             |
| 441         | 442          | 0.01         | 24           | 16.6      | 2893    | PASS             |
| 442         | 198          | 50           | 100          | 60.7      | 17396   | PASS             |
| 443         | 442          | 15           | 24           | 18.8      | 3270    | PASS             |