

Data Path : Z:\HPCHEM1\BNA E\DATA\BE093015\
 Data File : BE090985.D
 Acq On : 30 Sep 2015 18:58
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Oct 01 04:21:10 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE092815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 30 16:01:03 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	71	0.00
2	1,4-Dioxane	0.888	2.287	-157.5#	213#	0.00
3	n-Nitrosodimethylamine	0.879	0.766	12.9	69	0.00
4 S	2-Fluorophenol	1.045	0.937	10.3	73	0.00
5 S	Phenol-d6	1.280	1.160	9.4	72	0.01
6	bis(2-Chloroethyl)ether	1.524	1.604	-5.2	78	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	75	0.00
8 S	Nitrobenzene-d5	0.270	0.242	10.4	71	0.00
9	Nitrobenzene	0.289	0.231	20.1	67	0.00
10	Naphthalene	1.075	1.020	5.1	74	0.00
11	Hexachlorobutadiene	0.179	0.162	9.5	68	0.00
12	2-Methylnaphthalene	0.634	0.597	5.8	75	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	77	0.00
14 S	2,4,6-Tribromophenol	0.158	0.141	10.8	92	0.00
15 S	2-Fluorobiphenyl	1.355	1.169	13.7	72	0.00
16	Acenaphthylene	8.351	7.590	9.1	77	0.00
17	Acenaphthene	1.334	1.281	4.0	79	0.00
18	Fluorene	1.649	1.585	3.9	80	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	83	0.00
20	4-Bromophenyl-phenylether	0.189	0.168	11.1	80	0.00
21	Hexachlorobenzene	0.970	0.899	7.3	80	0.00
22	Pentachlorophenol	0.070	0.049	30.0#	91	0.00
23	Phenanthrene	1.265	1.162	8.1	84	-0.02
24	Anthracene	1.131	0.976	13.7	81	-0.02
25	Fluoranthene	1.503	1.301	13.4	88	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	87	0.00
27	Benzydine	0.090	0.102	-13.3	121	0.00
28	Pvrene	1.277	1.092	14.5	89	0.00
29 S	Terphenyl-d14	0.631	0.554	12.2	88	0.00
30	Benzo(a)anthracene	1.176	1.081	8.1	89	0.00
31	3,3'-Dichlorobenzidine	0.201	0.000	100.0#	0#	-20.97#
32	Chrysene	1.505	1.275	15.3	86	0.00
33	Bis(2-ethylhexyl)phthalate	0.469	0.332	29.2#	97	0.00
34	Indeno(1,2,3-cd)pyrene	1.212	1.143	5.7	95	0.00
35 I	Perylene-d12	1.000	1.000	0.0	89	0.00
36	Benzo(b)fluoranthene	1.102	1.012	8.2	92	0.00
37	Benzo(k)fluoranthene	1.515	1.276	15.8	84	0.00
38 C	Benzo(a)pyrene	1.022	0.990	3.1	100	0.00
39	Dibenzo(a,h)anthracene	1.009	0.954	5.5	100	0.00
40	Benzo(a,h,i)perylene	1.096	1.042	4.9	98	0.00

Data Path : Z:\HPCHEM1\BNA E\DATA\BE093015\
Data File : BE090985.D
Acq On : 30 Sep 2015 18:58
Operator : UM/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC001

Quant Time: Oct 01 04:21:10 2015
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE092815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 30 16:01:03 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				