

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE093016\
 Data File : BE092432.D
 Acq On : 30 Sep 2016 11:30
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 30 13:45:39 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE092216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 22 18:25:56 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	110	-0.02
2	1,4-Dioxane	0.400	0.361	9.8	107	-0.01
3	n-Nitrosodimethylamine	0.400	0.432	-8.0	103	0.00
4 S	2-Fluorophenol	0.400	0.372	7.0	91	0.00
5 S	Phenol-d6	0.400	0.320	20.0	91	0.00
6	bis(2-Chloroethyl)ether	0.400	0.361	9.8	105	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	104	0.00
8 S	Nitrobenzene-d5	0.400	0.409	-2.2	101	0.00
9	Naphthalene	0.400	0.397	0.8	104	-0.02
10	Hexachlorobutadiene	0.400	0.395	1.3	104	-0.02
11	2-Methylnaphthalene	0.400	0.386	3.5	102	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	104	0.00
13 S	2,4,6-Tribromophenol	0.400	0.384	4.0	85	0.00
14 S	2-Fluorobiphenyl	0.400	0.391	2.3	103	0.00
15	Acenaphthylene	0.400	0.366	8.5	98	0.00
16	Acenaphthene	0.400	0.392	2.0	98	0.00
17	Fluorene	0.400	0.373	6.8	98	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	94	0.02
19	Hexachlorobenzene	0.400	0.429	-7.2	108	0.00
20	Pentachlorophenol	0.400	0.351	12.3	78	0.00
21	Phenanthrene	0.400	0.420	-5.0	104	0.00
22	Anthracene	0.400	0.374	6.5	96	0.02
23	Fluoranthene	0.400	0.360	10.0	92	0.00
24 I	Chrysene-d12	5.000	5.000	0.0	95	0.00
25	Pyrene	0.400	0.377	5.8	93	0.00
26 S	Terphenyl-d14	0.400	0.375	6.3	90	0.02
27	Benzo(a)anthracene	0.400	0.375	6.3	91	0.00
28	Chrysene	0.400	0.434	-8.5	96	0.00
29	Bis(2-ethylhexyl)phthalate	0.400	0.284	29.0#	66	0.00
30	Indeno(1,2,3-cd)pyrene	0.400	0.342	14.5	82	0.04
31 I	Perylene-d12	5.000	5.000	0.0	89	0.04
32	Benzo(b)fluoranthene	0.400	0.360	10.0	80	0.02
33	Benzo(k)fluoranthene	0.400	0.403	-0.8	96	0.02
34 C	Benzo(a)pyrene	0.400	0.372	7.0	87	0.02
35	Dibenzo(a,h)anthracene	0.400	0.359	10.3	83	0.05
36	Benzo(g,h,i)perylene	0.400	0.373	6.8	85	0.05

(#) = Out of Range

SPCC's out = 0 CCC's out = 0