

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE093016\
 Data File : BE092432.D
 Acq On : 30 Sep 2016 11:30
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 30 13:45:39 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE092216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 22 18:25:56 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	-0.02
2	1,4-Dioxane	0.635	0.625	1.6	107	-0.01
3	n-Nitrosodimethylamine	0.588	0.497	15.5	103	0.00
4 S	2-Fluorophenol	0.999	0.775	22.4	91	0.00
5 S	Phenol-d6	1.388	0.999	28.0#	91	0.00
6	bis(2-Chloroethyl)ether	1.222	0.910	25.5#	105	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	104	0.00
8 S	Nitrobenzene-d5	0.299	0.244	18.4	101	0.00
9	Naphthalene	1.082	1.075	0.6	104	-0.02
10	Hexachlorobutadiene	0.170	0.168	1.2	104	-0.02
11	2-Methylnaphthalene	0.716	0.691	3.5	102	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	104	0.00
13 S	2,4,6-Tribromophenol	0.144	0.101	29.9#	85	0.00
14 S	2-Fluorobiphenyl	1.184	1.158	2.2	103	0.00
15	Acenaphthylene	7.240	6.634	8.4	98	0.00
16	Acenaphthene	1.134	1.110	2.1	98	0.00
17	Fluorene	1.438	1.340	6.8	98	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	94	0.02
19	Hexachlorobenzene	0.214	0.230	-7.5	108	0.00
20	Pentachlorophenol	0.050	0.035	30.0#	78	0.00
21	Phenanthrene	1.149	1.207	-5.0	104	0.00
22	Anthracene	1.105	1.033	6.5	96	0.02
23	Fluoranthene	5.460	4.920	9.9	92	0.00
24 I	Chrysene-d12	1.000	1.000	0.0	95	0.00
25	Pyrene	4.333	4.081	5.8	93	0.00
26 S	Terphenyl-d14	0.620	0.581	6.3	90	0.02
27	Benzo(a)anthracene	4.873	4.570	6.2	91	0.00
28	Chrysene	4.462	4.840	-8.5	96	0.00
29	Bis(2-ethylhexyl)phthalate	0.420	0.297	29.3#	66	0.00
30	Indeno(1,2,3-cd)pyrene	4.734	4.046	14.5	82	0.04
31 I	Perylene-d12	1.000	1.000	0.0	89	0.04
32	Benzo(b)fluoranthene	1.211	1.090	10.0	80	0.02
33	Benzo(k)fluoranthene	1.267	1.278	-0.9	96	0.02
34 C	Benzo(a)pyrene	1.187	1.104	7.0	87	0.02
35	Dibenzo(a,h)anthracene	1.088	0.976	10.3	83	0.05
36	Benzo(g,h,i)perylene	4.673	4.357	6.8	85	0.05

(#) = Out of Range

SPCC's out = 0 CCC's out = 0