

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE093016\
 Data File : BE092441.D
 Acq On : 30 Sep 2016 19:04
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Sep 30 19:42:02 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE092216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 22 18:25:56 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	104	-0.02
2	1,4-Dioxane	0.400	0.367	8.3	102	0.00
3	n-Nitrosodimethylamine	0.400	0.434	-8.5	97	0.00
4 S	2-Fluorophenol	0.400	0.368	8.0	85	0.00
5 S	Phenol-d6	0.400	0.312	22.0	83	0.00
6	bis(2-Chloroethyl)ether	0.400	0.342	14.5	92	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	101	0.00
8 S	Nitrobenzene-d5	0.400	0.400	0.0	95	0.00
9	Naphthalene	0.400	0.397	0.8	101	-0.02
10	Hexachlorobutadiene	0.400	0.383	4.3	98	-0.02
11	2-Methylnaphthalene	0.400	0.386	3.5	99	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	97	0.00
13 S	2,4,6-Tribromophenol	0.400	0.388	3.0	80	0.00
14 S	2-Fluorobiphenyl	0.400	0.406	-1.5	99	0.00
15	Acenaphthylene	0.400	0.386	3.5	96	0.00
16	Acenaphthene	0.400	0.404	-1.0	94	0.00
17	Fluorene	0.400	0.388	3.0	95	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	89	0.02
19	Hexachlorobenzene	0.400	0.417	-4.2	100	0.00
20	Pentachlorophenol	0.400	0.360	10.0	78	0.00
21	Phenanthrene	0.400	0.431	-7.7	102	0.00
22	Anthracene	0.400	0.381	4.8	94	0.02
23	Fluoranthene	0.400	0.384	4.0	93	0.00
24 I	Chrysene-d12	5.000	5.000	0.0	93	0.00
25	Pyrene	0.400	0.391	2.3	94	0.02
26 S	Terphenyl-d14	0.400	0.380	5.0	89	0.02
27	Benzo(a)anthracene	0.400	0.372	7.0	88	0.00
28	Chrysene	0.400	0.438	-9.5	95	0.00
29	Bis(2-ethylhexyl)phthalate	0.400	0.255	36.3#	59	0.00
30	Indeno(1,2,3-cd)pyrene	0.400	0.340	15.0	80	0.05
31 I	Perylene-d12	5.000	5.000	0.0	87	0.04
32	Benzo(b)fluoranthene	0.400	0.414	-3.5	91	0.02
33	Benzo(k)fluoranthene	0.400	0.354	11.5	83	0.04
34 C	Benzo(a)pyrene	0.400	0.370	7.5	85	0.02
35	Dibenzo(a,h)anthracene	0.400	0.361	9.8	82	0.05
36	Benzo(g,h,i)perylene	0.400	0.371	7.3	84	0.05

(#) = Out of Range

SPCC's out = 0 CCC's out = 0