

Data Path : Z:\HPCHEM1\BNA E\Data\BE100317\
 Data File : BE094159.D
 Acq On : 3 Oct 2017 17:55
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Quant Time: Oct 04 13:06:20 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE092517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 29 13:57:50 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	130	0.00
2	1,4-Dioxane	0.780	0.720	7.7	116	0.00
3	n-Nitrosodimethylamine	0.996	0.914	8.2	120	0.00
4 S	2-Fluorophenol	1.516	1.319	13.0	112	0.00
5 S	Phenol-d6	2.106	1.907	9.4	117	0.00
6	bis(2-Chloroethyl)ether	1.632	1.528	6.4	119	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	137	0.00
8 S	Nitrobenzene-d5	0.387	0.311	19.6	117	0.00
9	Naphthalene	4.516	4.502	0.3	138	0.00
10	Hexachlorobutadiene	0.169	0.167	1.2	142	0.00
11 SURR	2-Methylnaphthalene-d10	0.616	0.618	-0.3	140	0.00
12	2-Methylnaphthalene	0.731	0.736	-0.7	142	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	142	0.00
14 S	2,4,6-Tribromophenol	0.256	0.131	48.8#	85	0.00
15 S	2-Fluorobiphenyl	1.322	1.364	-3.2	140	0.00
16	Acenaphthylene	2.005	1.980	1.2	136	0.00
17	Acenaphthene	1.253	1.341	-7.0	141	0.00
18	Fluorene	1.700	1.734	-2.0	140	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	115	0.00
20	4-Bromophenyl-phenylether	0.164	0.199	-21.3	143	0.00
21	Hexachlorobenzene	0.104	0.162	-55.8#	182#	0.00
22	Pentachlorophenol	0.095	0.044	53.7#	54	0.00
23	Phenanthrene	1.023	1.229	-20.1	144	0.00
24	Anthracene	1.074	1.031	4.0	117	0.00
25 SURR	Fluoranthene-d10	3.873	3.641	6.0	113	0.00
26	Fluoranthene	1.147	1.103	3.8	114	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	104	0.00
28	Pvrene	1.267	1.302	-2.8	111	0.00
29 S	Terphenyl-d14	0.730	0.708	3.0	106	0.00
30	Benzo(a)anthracene	1.347	1.273	5.5	98	0.00
31	Chrysene	1.300	1.263	2.8	101	0.00
32	Bis(2-ethylhexyl)phthalate	6.447	2.897	55.1#	63	0.00
33	Indeno(1,2,3-cd)pyrene	1.391	1.291	7.2	100	0.00
34 I	Perylene-d12	1.000	1.000	0.0	100	-0.01
35	Benzo(b)fluoranthene	1.443	1.331	7.8	103	0.00
36	Benzo(k)fluoranthene	1.382	1.325	4.1	96	0.00
37 C	Benzo(a)pyrene	1.304	1.254	3.8	100	0.00
38	Dibenzo(a,h)anthracene	1.235	1.149	7.0	98	0.00
39	Benzo(a,h,i)perylene	1.222	1.183	3.2	101	0.00

Data Path : Z:\HPCHEM1\BNA E\Data\BE100317\
Data File : BE094159.D
Acq On : 3 Oct 2017 17:55
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Oct 04 13:06:20 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE092517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 29 13:57:50 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			