

Data Path : Z:\HPCHEM1\BNA E\Data\BE100317\
 Data File : BE094159.D
 Acq On : 3 Oct 2017 17:55
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Quant Time: Oct 04 13:06:20 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE092517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 29 13:57:50 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	130	0.00
2	1,4-Dioxane	0.400	0.370	7.5	116	0.00
3	n-Nitrosodimethylamine	0.400	0.367	8.3	120	0.00
4 S	2-Fluorophenol	0.400	0.348	13.0	112	0.00
5 S	Phenol-d6	0.400	0.362	9.5	117	0.00
6	bis(2-Chloroethyl)ether	0.400	0.374	6.5	119	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	137	0.00
8 S	Nitrobenzene-d5	0.400	0.322	19.5	117	0.00
9	Naphthalene	0.400	0.399	0.3	138	0.00
10	Hexachlorobutadiene	0.400	0.395	1.3	142	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.401	-0.3	140	0.00
12	2-Methylnaphthalene	0.400	0.402	-0.5	142	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	142	0.00
14 S	2,4,6-Tribromophenol	0.400	0.210	47.5#	85	0.00
15 S	2-Fluorobiphenyl	0.400	0.413	-3.2	140	0.00
16	Acenaphthylene	0.400	0.395	1.3	136	0.00
17	Acenaphthene	0.400	0.428	-7.0	141	0.00
18	Fluorene	0.400	0.408	-2.0	140	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	115	0.00
20	4-Bromophenyl-phenylether	0.400	0.485	-21.2	143	0.00
21	Hexachlorobenzene	0.400	0.626	-56.5#	182	0.00
22	Pentachlorophenol	0.400	0.184	54.0#	54	0.00
23	Phenanthrene	0.400	0.480	-20.0	144	0.00
24	Anthracene	0.400	0.384	4.0	117	0.00
25 SURR	Fluoranthene-d10	0.400	0.376	6.0	113	0.00
26	Fluoranthene	0.400	0.385	3.8	114	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	104	0.00
28	Pvrene	0.400	0.411	-2.7	111	0.00
29 S	Terphenyl-d14	0.400	0.388	3.0	106	0.00
30	Benzo(a)anthracene	0.400	0.378	5.5	98	0.00
31	Chrysene	0.400	0.389	2.8	101	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.088	78.0#	63	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.371	7.3	100	0.00
34 I	Perylene-d12	0.400	0.400	0.0	100	-0.01
35	Benzo(b)fluoranthene	0.400	0.369	7.8	103	0.00
36	Benzo(k)fluoranthene	0.400	0.383	4.3	96	0.00
37 C	Benzo(a)pyrene	0.400	0.385	3.8	100	0.00
38	Dibenzo(a,h)anthracene	0.400	0.372	7.0	98	0.00
39	Benzo(a,h,i)perylene	0.400	0.387	3.3	101	0.00

Data Path : Z:\HPCHEM1\BNA E\Data\BE100317\
Data File : BE094159.D
Acq On : 3 Oct 2017 17:55
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Oct 04 13:06:20 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE092517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 29 13:57:50 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				