

Data Path : Z:\HPCHEM1\BNA E\Data\BE100417\
 Data File : BE094170.D
 Acq On : 4 Oct 2017 17:08
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 04 20:21:15 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE092517.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 29 13:57:50 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	0.00
2	1,4-Dioxane	0.780	0.725	7.1	99	0.00
3	n-Nitrosodimethylamine	0.996	0.738	25.9#	82	0.00
4 S	2-Fluorophenol	1.516	1.338	11.7	96	0.00
5 S	Phenol-d6	2.106	2.101	0.2	109	0.00
6	bis(2-Chloroethyl)ether	1.632	1.516	7.1	100	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	127	0.00
8 S	Nitrobenzene-d5	0.387	0.309	20.2	107	0.00
9	Naphthalene	4.516	4.542	-0.6	129	0.00
10	Hexachlorobutadiene	0.169	0.159	5.9	125	0.00
11 SURR	2-Methylnaphthalene-d10	0.616	0.515	16.4	108	0.00
12	2-Methylnaphthalene	0.731	0.615	15.9	109	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	115	0.00
14 S	2,4,6-Tribromophenol	0.256	0.153	40.2#	80	0.00
15 S	2-Fluorobiphenyl	1.322	1.289	2.5	108	0.00
16	Acenaphthylene	2.005	1.980	1.2	111	0.00
17	Acenaphthene	1.253	1.326	-5.8	114	0.00
18	Fluorene	1.700	1.764	-3.8	116	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	90	0.00
20	4-Bromophenyl-phenylether	0.164	0.215	-31.1#	121	0.00
21	Hexachlorobenzene	0.104	0.187	-79.8#	164#	0.00
22	Pentachlorophenol	0.095	0.054	43.2#	53	0.00
23	Phenanthrene	1.023	1.331	-30.1#	122	0.00
24	Anthracene	1.074	1.051	2.1	93	0.00
25 SURR	Fluoranthene-d10	3.873	3.812	1.6	93	0.00
26	Fluoranthene	1.147	1.171	-2.1	94	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	81	0.00
28	Pvrene	1.267	1.369	-8.1	91	0.00
29 S	Terphenyl-d14	0.730	0.748	-2.5	87	0.00
30	Benzo(a)anthracene	1.347	1.298	3.6	78	0.00
31	Chrysene	1.300	1.274	2.0	79	0.00
32	Bis(2-ethylhexyl)phthalate	6.447	2.901	55.0#	49#	0.00
33	Indeno(1,2,3-cd)pyrene	1.391	1.304	6.3	79	0.00
34 I	Perylene-d12	1.000	1.000	0.0	79	0.00
35	Benzo(b)fluoranthene	1.443	1.281	11.2	78	0.00
36	Benzo(k)fluoranthene	1.382	1.346	2.6	77	0.00
37 C	Benzo(a)pyrene	1.304	1.274	2.3	80	0.00
38	Dibenzo(a,h)anthracene	1.235	1.158	6.2	78	0.00
39	Benzo(a,h,i)perylene	1.222	1.182	3.3	79	0.00

Data Path : Z:\HPCHEM1\BNA E\Data\BE100417\
Data File : BE094170.D
Acq On : 4 Oct 2017 17:08
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 04 20:21:15 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE092517.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 29 13:57:50 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			