

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE100418\
 Data File : BE097720.D
 Acq On : 4 Oct 2018 9:05
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 05 06:49:16 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE100318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 03 12:17:19 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	100	0.00
2	1,4-Dioxane	0.400	0.374	6.5	100	0.00
3	n-Nitrosodimethylamine	0.400	0.338	15.5	89	0.00
4 S	2-Fluorophenol	0.400	0.317	20.8	84	0.00
5 S	Phenol-d6	0.400	0.327	18.3	84	0.00
6	bis(2-Chloroethyl)ether	0.400	0.388	3.0	101	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	101	0.00
8 S	Nitrobenzene-d5	0.400	0.435	-8.7	117	0.00
9	Naphthalene	0.400	0.382	4.5	98	0.00
10	Hexachlorobutadiene	0.400	0.393	1.8	102	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.377	5.8	97	0.00
12	2-Methylnaphthalene	0.400	0.384	4.0	101	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	102	0.00
14 S	2,4,6-Tribromophenol	0.400	0.326	18.5	103	0.00
15 S	2-Fluorobiphenyl	0.400	0.383	4.3	103	0.00
16	Acenaphthylene	0.400	0.359	10.3	96	0.00
17	Acenaphthene	0.400	0.380	5.0	102	0.00
18	Fluorene	0.400	0.373	6.8	99	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	98	0.00
20	4-Bromophenyl-phenylether	0.400	0.379	5.3	100	0.00
21	Hexachlorobenzene	0.400	0.391	2.3	101	0.00
22	Pentachlorophenol	0.400	0.332	17.0	93	0.00
23	Phenanthrene	0.400	0.379	5.3	98	0.00
24	Anthracene	0.400	0.363	9.3	98	0.00
25 SURR	Fluoranthene-d10	0.400	0.366	8.5	97	0.00
26	Fluoranthene	0.400	0.368	8.0	97	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	93	0.00
28	Pvrene	0.400	0.378	5.5	98	0.00
29 S	Terphenyl-d14	0.400	0.373	6.8	97	0.00
30	Benzo(a)anthracene	0.400	0.356	11.0	93	0.00
31	Chrysene	0.400	0.381	4.8	96	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.285	28.8#	77	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.368	8.0	94	0.00
34 I	Perylene-d12	0.400	0.400	0.0	98	0.00
35	Benzo(b)fluoranthene	0.400	0.340	15.0	94	0.00
36	Benzo(k)fluoranthene	0.400	0.356	11.0	97	0.00
37 C	Benzo(a)pyrene	0.400	0.334	16.5	92	0.00
38	Dibenzo(a,h)anthracene	0.400	0.344	14.0	94	0.00
39	Benzo(a,h,i)perylene	0.400	0.350	12.5	93	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE100418\
Data File : BE097720.D
Acq On : 4 Oct 2018 9:05
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 05 06:49:16 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE100318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 03 12:17:19 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				