

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE100418\
 Data File : BE097720.D
 Acq On : 4 Oct 2018 9:05
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 05 06:49:16 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE100318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 03 12:17:19 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.797	0.745	6.5	100	0.00
3	n-Nitrosodimethylamine	0.725	0.612	15.6	89	0.00
4 S	2-Fluorophenol	0.946	0.751	20.6	84	0.00
5 S	Phenol-d6	1.450	1.185	18.3	84	0.00
6	bis(2-Chloroethyl)ether	1.501	1.457	2.9	101	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	101	0.00
8 S	Nitrobenzene-d5	0.130	0.141	-8.5	117	0.00
9	Naphthalene	4.031	3.853	4.4	98	0.00
10	Hexachlorobutadiene	0.214	0.210	1.9	102	0.00
11 SURR	2-Methylnaphthalene-d10	0.636	0.600	5.7	97	0.00
12	2-Methylnaphthalene	0.645	0.620	3.9	101	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	102	0.00
14 S	2,4,6-Tribromophenol	0.072	0.059	18.1	103	0.00
15 S	2-Fluorobiphenyl	1.761	1.688	4.1	103	0.00
16	Acenaphthylene	1.771	1.588	10.3	96	0.00
17	Acenaphthene	1.135	1.079	4.9	102	0.00
18	Fluorene	1.507	1.405	6.8	99	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	98	0.00
20	4-Bromophenyl-phenylether	0.210	0.199	5.2	100	0.00
21	Hexachlorobenzene	0.236	0.231	2.1	101	0.00
22	Pentachlorophenol	0.037	0.031	16.2	93	0.00
23	Phenanthrene	1.031	0.977	5.2	98	0.00
24	Anthracene	0.890	0.808	9.2	98	0.00
25 SURR	Fluoranthene-d10	5.063	4.636	8.4	97	0.00
26	Fluoranthene	1.313	1.208	8.0	97	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	93	0.00
28	Pvrene	1.016	0.960	5.5	98	0.00
29 S	Terphenyl-d14	0.817	0.761	6.9	97	0.00
30	Benzo(a)anthracene	1.075	0.957	11.0	93	0.00
31	Chrysene	1.202	1.144	4.8	96	0.00
32	Bis(2-ethylhexyl)phthalate	0.787	0.561	28.7#	77	0.00
33	Indeno(1,2,3-cd)pyrene	1.044	0.961	8.0	94	0.00
34 I	Perylene-d12	1.000	1.000	0.0	98	0.00
35	Benzo(b)fluoranthene	1.098	0.933	15.0	94	0.00
36	Benzo(k)fluoranthene	1.217	1.083	11.0	97	0.00
37 C	Benzo(a)pyrene	0.994	0.830	16.5	92	0.00
38	Dibenzo(a,h)anthracene	0.979	0.842	14.0	94	0.00
39	Benzo(a,h,i)perylene	1.024	0.896	12.5	93	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE100418\
Data File : BE097720.D
Acq On : 4 Oct 2018 9:05
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 05 06:49:16 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE100318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 03 12:17:19 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			