

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE100418\
 Data File : BE097724.D
 Acq On : 4 Oct 2018 13:16
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 05 06:50:04 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE100318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 03 12:17:19 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	59	0.01
2	1,4-Dioxane	0.400	0.392	2.0	62	0.00
3	n-Nitrosodimethylamine	0.400	0.371	7.3	58	0.00
4 S	2-Fluorophenol	0.400	0.414	-3.5	64	0.01
5 S	Phenol-d6	0.400	0.390	2.5	59	0.01
6	bis(2-Chloroethyl)ether	0.400	0.392	2.0	60	0.02
7 I	Naphthalene-d8	0.400	0.400	0.0	61	0.01
8 S	Nitrobenzene-d5	0.400	0.512	-28.0#	83	0.01
9	Naphthalene	0.400	0.379	5.3	59	0.01
10	Hexachlorobutadiene	0.400	0.385	3.8	61	0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.379	5.3	59	0.01
12	2-Methylnaphthalene	0.400	0.378	5.5	60	0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	63	0.01
14 S	2,4,6-Tribromophenol	0.400	0.425	-6.2	83	0.00
15 S	2-Fluorobiphenyl	0.400	0.365	8.8	61	0.00
16	Acenaphthylene	0.400	0.357	10.8	59	0.01
17	Acenaphthene	0.400	0.379	5.3	63	0.01
18	Fluorene	0.400	0.383	4.3	63	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	62	0.00
20	4-Bromophenyl-phenylether	0.400	0.388	3.0	65	0.01
21	Hexachlorobenzene	0.400	0.381	4.8	62	0.01
22	Pentachlorophenol	0.400	0.408	-2.0	72	0.00
23	Phenanthrene	0.400	0.398	0.5	65	0.01
24	Anthracene	0.400	0.374	6.5	63	0.01
25 SURR	Fluoranthene-d10	0.400	0.374	6.5	62	0.00
26	Fluoranthene	0.400	0.373	6.8	62	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	58	0.00
28	Pvrene	0.400	0.389	2.8	63	0.00
29 S	Terphenyl-d14	0.400	0.392	2.0	64	0.00
30	Benzo(a)anthracene	0.400	0.357	10.8	58	0.00
31	Chrysene	0.400	0.394	1.5	62	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.347	13.3	58	0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.363	9.3	57	0.01
34 I	Perylene-d12	0.400	0.400	0.0	57	0.00
35	Benzo(b)fluoranthene	0.400	0.369	7.8	59	0.00
36	Benzo(k)fluoranthene	0.400	0.384	4.0	61	0.00
37 C	Benzo(a)pyrene	0.400	0.365	8.8	58	0.00
38	Dibenzo(a,h)anthracene	0.400	0.362	9.5	57	0.00
39	Benzo(a,h,i)perylene	0.400	0.375	6.3	58	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE100418\
Data File : BE097724.D
Acq On : 4 Oct 2018 13:16
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 05 06:50:04 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE100318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 03 12:17:19 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				