

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE100418\
 Data File : BE097743.D
 Acq On : 5 Oct 2018 2:37
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 05 16:17:08 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE100318.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 03 12:17:19 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	68	0.00
2	1,4-Dioxane	0.797	0.736	7.7	67	0.00
3	n-Nitrosodimethylamine	0.725	0.628	13.4	62	0.00
4 S	2-Fluorophenol	0.946	1.076	-13.7	81	0.00
5 S	Phenol-d6	1.450	1.831	-26.3#	87	0.00
6	bis(2-Chloroethyl)ether	1.501	1.413	5.9	66	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	68	0.00
8 S	Nitrobenzene-d5	0.130	0.203	-56.2#	114	0.00
9	Naphthalene	4.031	3.817	5.3	66	0.00
10	Hexachlorobutadiene	0.214	0.205	4.2	68	0.00
11 SURR	2-Methylnaphthalene-d10	0.636	0.600	5.7	66	0.00
12	2-Methylnaphthalene	0.645	0.610	5.4	67	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	72	0.00
14 S	2,4,6-Tribromophenol	0.072	0.096	-33.3#	118	-0.01
15 S	2-Fluorobiphenyl	1.761	1.596	9.4	69	-0.02
16	Acenaphthylene	1.771	1.600	9.7	68	0.00
17	Acenaphthene	1.135	1.062	6.4	71	0.00
18	Fluorene	1.507	1.380	8.4	69	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	71	0.00
20	4-Bromophenyl-phenylether	0.210	0.201	4.3	72	0.00
21	Hexachlorobenzene	0.236	0.215	8.9	68	0.00
22	Pentachlorophenol	0.037	0.049	-32.4#	107	-0.01
23	Phenanthrene	1.031	0.971	5.8	70	0.00
24	Anthracene	0.890	0.823	7.5	71	0.00
25 SURR	Fluoranthene-d10	5.063	4.767	5.8	71	0.00
26	Fluoranthene	1.313	1.223	6.9	70	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	66	0.00
28	Pvrene	1.016	0.989	2.7	71	0.00
29 S	Terphenyl-d14	0.817	0.786	3.8	71	0.00
30	Benzo(a)anthracene	1.075	1.058	1.6	73	-0.01
31	Chrysene	1.202	1.173	2.4	70	0.00
32	Bis(2-ethylhexyl)phthalate	0.787	0.911	-15.8	89	0.00
33	Indeno(1,2,3-cd)pyrene	1.044	0.985	5.7	68	0.00
34 I	Perylene-d12	1.000	1.000	0.0	71	0.00
35	Benzo(b)fluoranthene	1.098	0.967	11.9	71	0.00
36	Benzo(k)fluoranthene	1.217	1.131	7.1	74	0.00
37 C	Benzo(a)pyrene	0.994	0.895	10.0	72	0.00
38	Dibenzo(a,h)anthracene	0.979	0.833	14.9	68	-0.01
39	Benzo(a,h,i)perylene	1.024	0.885	13.6	67	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE100418\
Data File : BE097743.D
Acq On : 5 Oct 2018 2:37
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 05 16:17:08 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE100318.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 03 12:17:19 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			