

Data Path : Z:\HPCHEM1\BNA E\DATA\BE100617\
 Data File : BE094180.D
 Acq On : 6 Oct 2017 15:07
 Operator : SJ/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICSVBE100617

Quant Time: Oct 06 15:47:25 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE100617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 06 15:16:13 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	104	0.00
2	1,4-Dioxane	0.400	0.378	5.5	92	0.02
3	n-Nitrosodimethylamine	0.400	0.383	4.3	100	0.02
4 S	2-Fluorophenol	0.400	0.392	2.0	102	0.00
5 S	Phenol-d6	0.400	0.390	2.5	103	0.00
6	bis(2-Chloroethyl)ether	0.400	0.397	0.8	105	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	107	0.00
8 S	Nitrobenzene-d5	0.400	0.397	0.8	107	0.00
9	Naphthalene	0.400	0.405	-1.3	106	0.00
10	Hexachlorobutadiene	0.400	0.407	-1.7	108	-0.02
11 SURR	2-Methylnaphthalene-d10	0.400	0.405	-1.3	107	0.00
12	2-Methylnaphthalene	0.400	0.404	-1.0	107	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	109	0.00
14 S	2,4,6-Tribromophenol	0.400	0.328	18.0	95	0.00
15 S	2-Fluorobiphenyl	0.400	0.397	0.8	108	0.00
16	Acenaphthylene	0.400	0.391	2.3	106	0.00
17	Acenaphthene	0.400	0.398	0.5	106	0.00
18	Fluorene	0.400	0.394	1.5	107	-0.02
19 I	Phenanthrene-d10	0.400	0.400	0.0	105	0.00
20	4-Bromophenyl-phenylether	0.400	0.389	2.8	101	0.00
21	Hexachlorobenzene	0.400	0.373	6.8	99	0.00
22	Pentachlorophenol	0.400	0.343	14.2	89	0.00
23	Phenanthrene	0.400	0.389	2.8	101	0.00
24	Anthracene	0.400	0.409	-2.2	110	0.00
25 SURR	Fluoranthene-d10	0.400	0.375	6.3	97	0.00
26	Fluoranthene	0.400	0.376	6.0	98	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	95	0.00
28	Pvrene	0.400	0.399	0.3	95	0.00
29 S	Terphenyl-d14	0.400	0.398	0.5	95	0.00
30	Benzo(a)anthracene	0.400	0.392	2.0	93	0.00
31	Chrysene	0.400	0.406	-1.5	97	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.355	11.3	89	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.406	-1.5	96	0.00
34 I	Perylene-d12	0.400	0.400	0.0	95	0.00
35	Benzo(b)fluoranthene	0.400	0.404	-1.0	96	0.00
36	Benzo(k)fluoranthene	0.400	0.404	-1.0	96	0.00
37 C	Benzo(a)pyrene	0.400	0.397	0.8	95	0.00
38	Dibenzo(a,h)anthracene	0.400	0.397	0.8	94	0.00
39	Benzo(a,h,i)perylene	0.400	0.404	-1.0	97	-0.01

Data Path : Z:\HPCHEM1\BNA E\DATA\BE100617\
Data File : BE094180.D
Acq On : 6 Oct 2017 15:07
Operator : SJ/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE100617

Quant Time: Oct 06 15:47:25 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE100617.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 06 15:16:13 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				