

Data Path : Z:\HPCHEM1\BNA E\Data\BE100617\
 Data File : BE094182.D
 Acq On : 6 Oct 2017 16:51
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 07 01:32:03 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE100617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 06 15:16:13 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	115	0.00
2	1,4-Dioxane	0.725	0.730	-0.7	108	0.01
3	n-Nitrosodimethylamine	0.777	0.756	2.7	112	0.01
4 S	2-Fluorophenol	1.333	1.289	3.3	112	0.00
5 S	Phenol-d6	1.932	1.811	6.3	109	0.00
6	bis(2-Chloroethyl)ether	1.569	1.500	4.4	112	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	116	0.00
8 S	Nitrobenzene-d5	0.333	0.324	2.7	114	0.00
9	Naphthalene	4.505	4.525	-0.4	114	0.00
10	Hexachlorobutadiene	0.171	0.173	-1.2	117	-0.02
11 SURR	2-Methylnaphthalene-d10	0.616	0.613	0.5	114	0.00
12	2-Methylnaphthalene	0.740	0.740	0.0	115	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	119	0.00
14 S	2,4,6-Tribromophenol	0.174	0.147	15.5	108	0.00
15 S	2-Fluorobiphenyl	1.370	1.355	1.1	118	0.00
16	Acenaphthylene	2.026	1.977	2.4	116	0.00
17	Acenaphthene	1.338	1.330	0.6	116	0.00
18	Fluorene	1.753	1.698	3.1	115	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	117	0.00
20	4-Bromophenyl-phenylether	0.230	0.214	7.0	107	0.00
21	Hexachlorobenzene	0.207	0.185	10.6	105	0.00
22	Pentachlorophenol	0.056	0.045	19.6	92	0.00
23	Phenanthrene	1.386	1.308	5.6	109	0.00
24	Anthracene	1.057	1.058	-0.1	119	0.00
25 SURR	Fluoranthene-d10	3.995	3.639	8.9	105	0.00
26	Fluoranthene	1.216	1.100	9.5	105	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	105	0.00
28	Pvrene	1.376	1.355	1.5	103	0.00
29 S	Terphenyl-d14	0.755	0.737	2.4	103	0.00
30	Benzo(a)anthracene	1.314	1.275	3.0	101	0.00
31	Chrysene	1.297	1.296	0.1	105	0.00
32	Bis(2-ethylhexyl)phthalate	3.027	2.609	13.8	95	0.00
33	Indeno(1,2,3-cd)pyrene	1.287	1.291	-0.3	104	0.00
34 I	Perylene-d12	1.000	1.000	0.0	105	0.00
35	Benzo(b)fluoranthene	1.345	1.319	1.9	103	0.00
36	Benzo(k)fluoranthene	1.368	1.327	3.0	102	0.00
37 C	Benzo(a)pyrene	1.272	1.290	-1.4	108	0.00
38	Dibenzo(a,h)anthracene	1.179	1.147	2.7	102	0.00
39	Benzo(a,h,i)perylene	1.204	1.169	2.9	103	-0.01

Data Path : Z:\HPCHEM1\BNA E\Data\BE100617\
Data File : BE094182.D
Acq On : 6 Oct 2017 16:51
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 07 01:32:03 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE100617.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 06 15:16:13 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			