


Data Path : Z:\HPCHEM1\BNA E\Data\BE100617\
 Data File : BE094181.D
 Acq On : 6 Oct 2017 16:15
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE100617.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Oct 06 15:16:13 2017



AutoFind: Scans 2566, 2567, 2568; Background Corrected with Scan 2559

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	49.4	10602	PASS
68	69	0.00	2	0.9	92	PASS
69	198	0.00	100	48.1	10331	PASS
70	69	0.00	2	0.4	39	PASS
127	198	10	80	51.9	11138	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	21458	PASS
199	198	5	9	7.0	1509	PASS
275	198	10	60	23.7	5095	PASS
365	198	1	100	2.6	557	PASS
441	442	0.01	24	16.7	2243	PASS
442	198	50	100	62.6	13426	PASS
443	442	15	24	19.1	2562	PASS