

Data Path : Z:\HPCHEM1\BNA E\DATA\BE100715\
 Data File : BE090987.D
 Acq On : 7 Oct 2015 14:04
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Oct 08 01:44:51 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE092815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 07 17:11:43 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	176#	0.00
2	1,4-Dioxane	0.888	1.587	-78.7#	367#	0.00
3	n-Nitrosodimethylamine	0.879	0.561	36.2#	124	0.00
4 S	2-Fluorophenol	1.045	1.044	0.1	201#	0.00
5 S	Phenol-d6	1.280	1.226	4.2	189#	0.00
6	bis(2-Chloroethyl)ether	1.524	2.981	-95.6#	357#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	174#	0.00
8 S	Nitrobenzene-d5	0.270	0.362	-34.1#	247#	0.00
9	Nitrobenzene	0.289	0.396	-37.0#	266#	0.00
10	Naphthalene	1.075	1.019	5.2	173#	0.00
11	Hexachlorobutadiene	0.179	0.150	16.2	147	0.00
12	2-Methylnaphthalene	0.634	0.668	-5.4	195#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	169#	0.00
14 S	2,4,6-Tribromophenol	0.158	0.135	14.6	192#	0.00
15 S	2-Fluorobiphenyl	1.355	1.360	-0.4	182#	0.00
16	Acenaphthylene	8.351	8.074	3.3	178#	0.00
17	Acenaphthene	1.334	1.277	4.3	172#	0.00
18	Fluorene	1.649	1.630	1.2	179#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	177#	0.00
20	4-Bromophenyl-phenylether	0.189	0.185	2.1	188#	0.00
21	Hexachlorobenzene	0.970	0.202	79.2#	39#	0.00
22	Pentachlorophenol	0.070	0.045	35.7#	177#	0.00
23	Phenanthrene	1.265	1.164	8.0	179#	0.00
24	Anthracene	1.131	1.095	3.2	194#	0.00
25	Fluoranthene	1.503	1.391	7.5	201#	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	195#	0.00
27	Benzydine	0.090	0.156	-73.3#	415#	0.00
28	Pvrene	1.277	1.087	14.9	199#	0.00
29 S	Terphenyl-d14	0.631	0.621	1.6	221#	0.00
30	Benzo(a)anthracene	1.176	1.197	-1.8	221#	0.00
31	3,3'-Dichlorobenzidine	0.201	0.000	100.0#	0#	-20.84#
32	Chrysene	1.505	1.230	18.3	187#	0.00
33	Bis(2-ethylhexyl)phthalate	0.469	0.560	-19.4	367#	0.00
34	Indeno(1,2,3-cd)pyrene	1.212	1.426	-17.7	266#	0.00
35 I	Perylene-d12	1.000	1.000	0.0	239#	0.00
36	Benzo(b)fluoranthene	1.102	1.164	-5.6	281#	0.00
37	Benzo(k)fluoranthene	1.515	1.260	16.8	222#	0.00
38 C	Benzo(a)pyrene	1.022	1.122	-9.8	304#	0.00
39	Dibenzo(a,h)anthracene	1.009	1.282	-27.1#	361#	0.00
40	Benzo(a,h,i)perylene	1.096	1.323	-20.7	332#	0.00

Data Path : Z:\HPCHEM1\BNA E\DATA\BE100715\
Data File : BE090987.D
Acq On : 7 Oct 2015 14:04
Operator : UM/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC001

Quant Time: Oct 08 01:44:51 2015
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE092815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 07 17:11:43 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0