

Data Path : Z:\HPCHEM1\BNA E\DATA\BE100715\
 Data File : BE090994.D
 Acq On : 7 Oct 2015 19:59
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Oct 08 02:12:07 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE092815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 08 01:45:47 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	165	0.00
2	1,4-Dioxane	1.000	2.601	-160.1#	377	-0.03
3	n-Nitrosodimethylamine	1.000	0.577	42.3#	100	-0.06
4 S	2-Fluorophenol	1.000	1.083	-8.3	185	0.00
5 S	Phenol-d6	1.000	0.964	3.6	177	0.00
6	bis(2-Chloroethyl)ether	1.000	2.383	-138.3#	359	-0.01
7 I	Naphthalene-d8	5.000	5.000	0.0	164	0.00
8 S	Nitrobenzene-d5	1.000	1.236	-23.6	233	0.00
9	Nitrobenzene	1.000	1.292	-29.2#	255	0.00
10	Naphthalene	1.000	1.003	-0.3	165	0.00
11	Hexachlorobutadiene	1.000	0.854	14.6	139	0.00
12	2-Methylnaphthalene	1.000	1.058	-5.8	185	0.00
13 I	Acenaphthene-d10	5.000	5.000	0.0	163	0.00
14 S	2,4,6-Tribromophenol	1.000	0.926	7.4	184	0.00
15 S	2-Fluorobiphenyl	1.000	0.987	1.3	173	0.00
16	Acenaphthylene	1.000	0.951	4.9	170	0.00
17	Acenaphthene	1.000	0.939	6.1	163	0.00
18	Fluorene	1.000	0.981	1.9	172	0.00
19 I	Phenanthrene-d10	5.000	5.000	0.0	167	0.00
20	4-Bromophenyl-phenylether	1.000	1.001	-0.1	182	0.00
21	Hexachlorobenzene	1.000	0.126	87.4#	37	0.00
22	Pentachlorophenol	1.000	0.909	9.1	185	0.00
23	Phenanthrene	1.000	0.968	3.2	169	-0.01
24	Anthracene	1.000	1.018	-1.8	183	-0.01
25	Fluoranthene	1.000	1.042	-4.2	190	0.00
26 I	Chrysene-d12	5.000	5.000	0.0	182	0.00
27	Benzydine	1.000	1.219	-21.9	226	0.00
28	Pvrene	1.000	0.930	7.0	185	0.00
29 S	Terphenyl-d14	1.000	1.026	-2.6	205	0.00
30	Benzo(a)anthracene	1.000	1.033	-3.3	198	0.00
31	3,3'-Dichlorobenzidine	1.000	0.022	97.8#	0	-20.84#
32	Chrysene	1.000	0.906	9.4	172	0.00
33	Bis(2-ethylhexyl)phthalate	1.000	1.322	-32.2#	306	0.00
34	Indeno(1,2,3-cd)pyrene	1.000	1.177	-17.7	238	0.00
35 I	Perylene-d12	5.000	5.000	0.0	218	0.00
36	Benzo(b)fluoranthene	1.000	1.081	-8.1	262	0.00
37	Benzo(k)fluoranthene	1.000	0.815	18.5	190	0.00
38 C	Benzo(a)pyrene	1.000	1.079	-7.9	272	0.00
39	Dibenzo(a,h)anthracene	1.000	1.248	-24.8	323	0.00
40	Benzo(a,h,i)perylene	1.000	1.187	-18.7	298	0.00

Data Path : Z:\HPCHEM1\BNA E\DATA\BE100715\
Data File : BE090994.D
Acq On : 7 Oct 2015 19:59
Operator : UM/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC001

Quant Time: Oct 08 02:12:07 2015
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE092815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 08 01:45:47 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)
----------	--------	-------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0