

Data Path : Z:\HPCHEM1\BNA E\DATA\BE100716\
 Data File : BE092444.D
 Acq On : 7 Oct 2016 12:47
 Operator : UM/SJ
 Sample : H5095-02
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 OUTFALL001-161004

Quant Time: Oct 10 10:41:54 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE092216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 22 18:25:56 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.17	152	9268	5.00	ng	-0.02
7) Naphthalene-d8	10.95	136	41656	5.00	ng	0.00
12) Acenaphthene-d10	14.81	164	32121	5.00	ng	0.00
18) Phenanthrene-d10	17.58	188	67017	5.00	ng	0.02
24) Chrysene-d12	21.75	240	93209	5.00	ng	0.00
31) Perylene-d12	24.11	264	77385	5.00	ng	0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.66	112	6261	2.94	ng	-0.02
5) Phenol-d6	7.34	99	4439	1.49	ng	-0.02
8) Nitrobenzene-d5	9.34	82	9851	3.34	ng	-0.02
13) 2,4,6-Tribromophenol	16.31	330	7074	6.25	ng	-0.01
14) 2-Fluorobiphenyl	13.43	172	29749	3.91	ng	-0.01
26) Terphenyl-d14	20.17	244	53736	4.65	ng	0.00
Target Compounds						Qvalue
27) Benzo(a)anthracene	21.72	228	2275m	0.03	ng	
28) Chrysene	21.77	228	1907m	0.02	ng	
29) Bis(2-ethylhexyl)phthalate	21.63	149	4595	0.59	ng	# 76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE100716\
Data File : BE092444.D
Acq On : 7 Oct 2016 12:47
Operator : UM/SJ
Sample : H5095-02
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
OUTFALL001-161004

Quant Time: Oct 10 10:41:54 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE092216.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 22 18:25:56 2016
Response via : Initial Calibration

