

Data Path : Z:\HPCHEM1\BNA E\DATA\BE100716\
 Data File : BE092443.D
 Acq On : 7 Oct 2016 11:53
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 07 23:06:54 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE092216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 22 18:25:56 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	109	-0.02
2	1,4-Dioxane	0.400	0.389	2.8	113	-0.01
3	n-Nitrosodimethylamine	0.400	0.436	-9.0	103	0.00
4 S	2-Fluorophenol	0.400	0.389	2.8	97	0.00
5 S	Phenol-d6	0.400	0.320	20.0	91	0.00
6	bis(2-Chloroethyl)ether	0.400	0.345	13.8	98	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	104	0.00
8 S	Nitrobenzene-d5	0.400	0.408	-2.0	101	0.00
9	Naphthalene	0.400	0.398	0.5	104	-0.02
10	Hexachlorobutadiene	0.400	0.396	1.0	104	-0.02
11	2-Methylnaphthalene	0.400	0.390	2.5	103	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	102	0.00
13 S	2,4,6-Tribromophenol	0.400	0.385	3.8	83	0.00
14 S	2-Fluorobiphenyl	0.400	0.410	-2.5	106	0.00
15	Acenaphthylene	0.400	0.381	4.8	100	0.00
16	Acenaphthene	0.400	0.400	0.0	99	0.00
17	Fluorene	0.400	0.384	4.0	99	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	93	0.02
19	Hexachlorobenzene	0.400	0.428	-7.0	106	0.00
20	Pentachlorophenol	0.400	0.446	-11.5	119	0.00
21	Phenanthrene	0.400	0.432	-8.0	106	0.00
22	Anthracene	0.400	0.381	4.8	97	0.02
23	Fluoranthene	0.400	0.375	6.3	94	0.00
24 I	Chrysene-d12	5.000	5.000	0.0	94	0.00
25	Pyrene	0.400	0.393	1.8	96	0.00
26 S	Terphenyl-d14	0.400	0.403	-0.8	95	0.02
27	Benzo(a)anthracene	0.400	0.367	8.3	88	0.00
28	Chrysene	0.400	0.435	-8.7	95	0.00
29	Bis(2-ethylhexyl)phthalate	0.400	0.310	22.5	72	0.00
30	Indeno(1,2,3-cd)pyrene	0.400	0.339	15.3	81	0.03
31 I	Perylene-d12	5.000	5.000	0.0	88	0.02
32	Benzo(b)fluoranthene	0.400	0.373	6.8	82	0.02
33	Benzo(k)fluoranthene	0.400	0.388	3.0	92	0.02
34 C	Benzo(a)pyrene	0.400	0.363	9.3	85	0.02
35	Dibenzo(a,h)anthracene	0.400	0.351	12.3	81	0.05
36	Benzo(g,h,i)perylene	0.400	0.371	7.3	84	0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0