

Data Path : Z:\HPCHEM1\BNA E\DATA\BE100815\
 Data File : BE090998.D
 Acq On : 8 Oct 2015 11:57
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Oct 09 03:17:49 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE092815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 07 17:11:43 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	167#	0.00
2	1,4-Dioxane	0.888	1.833	-106.4#	402#	-0.10
3	n-Nitrosodimethylamine	0.879	0.680	22.6	143	-0.13
4 S	2-Fluorophenol	1.045	1.136	-8.7	207#	-0.02
5 S	Phenol-d6	1.280	1.471	-14.9	215#	-0.02
6	bis(2-Chloroethyl)ether	1.524	3.741	-145.5#	425#	-0.05
7 I	Naphthalene-d8	1.000	1.000	0.0	172#	0.00
8 S	Nitrobenzene-d5	0.270	0.370	-37.0#	249#	-0.02
9	Nitrobenzene	0.289	0.430	-48.8#	286#	-0.01
10	Naphthalene	1.075	1.027	4.5	172#	-0.02
11	Hexachlorobutadiene	0.179	0.148	17.3	143	-0.02
12	2-Methylnaphthalene	0.634	0.687	-8.4	198#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	176#	0.00
14 S	2,4,6-Tribromophenol	0.158	0.161	-1.9	240#	0.00
15 S	2-Fluorobiphenyl	1.355	1.323	2.4	185#	-0.01
16	Acenaphthylene	8.351	8.262	1.1	190#	0.00
17	Acenaphthene	1.334	1.266	5.1	178#	0.00
18	Fluorene	1.649	1.623	1.6	186#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	185#	-0.01
20	4-Bromophenyl-phenylether	0.189	0.182	3.7	194#	-0.01
21	Hexachlorobenzene	0.970	0.198	79.6#	40#	-0.01
22	Pentachlorophenol	0.070	0.073	-4.3	302#	0.00
23	Phenanthrene	1.265	1.158	8.5	186#	-0.01
24	Anthracene	1.131	1.099	2.8	204#	-0.01
25	Fluoranthene	1.503	1.375	8.5	208#	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	195#	0.00
27	Benzydine	0.090	0.208	-131.1#	552#	-0.03
28	Pvrene	1.277	1.134	11.2	207#	0.00
29 S	Terphenyl-d14	0.631	0.625	1.0	221#	0.00
30	Benzo(a)anthracene	1.176	1.201	-2.1	221#	0.00
31	3,3'-Dichlorobenzidine	0.201	0.000	100.0#	0#	-20.84#
32	Chrysene	1.505	1.240	17.6	188#	0.00
33	Bis(2-ethylhexyl)phthalate	0.469	0.576	-22.8	375#	0.00
34	Indeno(1,2,3-cd)pyrene	1.212	1.292	-6.6	240#	0.00
35 I	Perylene-d12	1.000	1.000	0.0	229#	0.00
36	Benzo(b)fluoranthene	1.102	1.241	-12.6	288#	0.00
37	Benzo(k)fluoranthene	1.515	1.156	23.7	195#	0.00
38 C	Benzo(a)pyrene	1.022	1.112	-8.8	289#	0.00
39	Dibenzo(a,h)anthracene	1.009	1.212	-20.1	327#	0.00
40	Benzo(a,h,i)perylene	1.096	1.248	-13.9	301#	0.00

Data Path : Z:\HPCHEM1\BNA E\DATA\BE100815\
Data File : BE090998.D
Acq On : 8 Oct 2015 11:57
Operator : UM/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC001

Quant Time: Oct 09 03:17:49 2015
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE092815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 07 17:11:43 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				