

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA E\DATA\BE101117\
 Data File : BE094302.D
 Acq On : 12 Oct 2017 5:06
 Operator : SJ/JU
 Sample : I5550-03
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 C0AF3

Manual Integrations
 APPROVED

Sohil
 10/13/2017 12:06:08 PM

Quant Time: Oct 12 09:33:55 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE092017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 12 08:55:52 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.92	152	1588	0.40	ng/ul	0.00
2) Naphthalene-d8	10.71	136	8202	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.54	164	4812	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.27	188	10180m	0.40	ng/ul	0.02
16) Chrysene-d12	21.44	240	9495	0.40	ng/ul	0.00
20) Perylene-d12	23.96	264	9572m	0.40	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Methylnaphthalene-d10	12.29	152	4874	0.38	ng/ul	0.00
14) Fluoranthene-d10	19.29	212	11908	0.35	ng/ul	0.00

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA E\DATA\BE101117\
Data File : BE094302.D
Acq On : 12 Oct 2017 5:06
Operator : SJ/JU
Sample : I5550-03
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
C0AF3

Manual Integrations
APPROVED

Sohil
10/13/2017 12:06:08 PM

Quant Time: Oct 12 09:33:55 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE092017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 12 08:55:52 2017
Response via : Initial Calibration

