

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE101317\
 Data File : BE094334.D
 Acq On : 13 Oct 2017 21:50
 Operator : SJ/JU
 Sample : I5568-05DL 5X
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 B-75(1-3)-092917DL

Quant Time: Oct 14 04:25:27 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE092017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Oct 14 03:57:24 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.93	152	4118	0.40	ng/ul	0.00
2) Naphthalene-d8	10.72	136	9645	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.54	164	4900	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.27	188	9585	0.40	ng/ul	0.00
16) Chrysene-d12	21.45	240	8570	0.40	ng/ul	0.00
20) Perylene-d12	24.00	264	9498	0.40	ng/ul	0.02
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.30	152	1475	0.10	ng/ul	0.00
14) Fluoranthene-d10	19.30	212	1782	0.06	ng/ul	0.00
Target Compounds						
5) 2-Methylnaphthalene	12.37	142	1225	0.08	ng/ul	92
8) Acenaphthene	14.60	153	3508	0.22	ng/ul#	84
9) Fluorene	15.59	166	2864	0.16	ng/ul#	67
17) Pyrene	19.69	202	4413	0.19	ng/ul#	60

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE101317\
Data File : BE094334.D
Acq On : 13 Oct 2017 21:50
Operator : SJ/JU
Sample : I5568-05DL 5X
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
B-75(1-3)-092917DL

Quant Time: Oct 14 04:25:27 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE092017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Oct 14 03:57:24 2017
Response via : Initial Calibration

