

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA E\DATA\BE101417\
 Data File : BE094359.D
 Acq On : 14 Oct 2017 17:32
 Operator : SJ/JU
 Sample : I5522-14
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 C0AE3

Manual Integrations
 APPROVED

Sohil
 10/16/2017 8:25:06 PM

Quant Time: Oct 15 01:50:34 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE092017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Oct 14 05:22:31 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.91	152	1282	0.40	ng/ul	-0.01
2) Naphthalene-d8	10.70	136	6764	0.40	ng/ul	-0.01
6) Acenaphthene-d10	14.52	164	4499	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.26	188	10931m	0.40	ng/ul	-0.02
16) Chrysene-d12	21.42	240	10786	0.40	ng/ul	-0.01
20) Perylene-d12	23.94	264	9809m	0.40	ng/ul	-0.03
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.28	152	3179	0.30	ng/ul	0.00
14) Fluoranthene-d10	19.28	212	10449	0.29	ng/ul	-0.01
Target Compounds				Ovalue		
3) Naphthalene	10.74	128	2270	0.137	ng/ul#	94
5) 2-Methylnaphthalene	12.36	142	482	0.045	ng/ul	95
7) Acenaphthylene	14.24	152	602	0.034	ng/ul#	73
9) Fluorene	15.57	166	389	0.023	ng/ul#	88
12) Phenanthrene	17.29	178	5340	0.187	ng/ul#	89
13) Anthracene	17.38	178	1259	0.045	ng/ul#	81
15) Fluoranthene	19.31	202	14797	0.394	ng/ul	84
17) Pyrene	19.67	202	15012	0.500	ng/ul#	82
18) Benzo(a)anthracene	21.40	228	7582	0.265	ng/ul#	89
19) Chrysene	21.46	228	8654	0.262	ng/ul	94
21) Benzo(b)fluoranthene	23.17	252	10457m	0.392	ng/ul	
22) Benzo(k)fluoranthene	23.21	252	3726m	0.121	ng/ul	
23) Benzo(a)pyrene	23.83	252	7418m	0.268	ng/ul	
24) Indeno(1,2,3-cd)pyrene	26.63	276	4980m	0.176	ng/ul	
26) Benzo(g,h,i)perylene	27.46	276	4857m	0.199	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA E\DATA\BE101417\
Data File : BE094359.D
Acq On : 14 Oct 2017 17:32
Operator : SJ/JU
Sample : I5522-14
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
C0AE3

Manual Integrations
APPROVED

Sohil
10/16/2017 8:25:06 PM

Quant Time: Oct 15 01:50:34 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE092017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Oct 14 05:22:31 2017
Response via : Initial Calibration

