

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA E\DATA\BE101417\
 Data File : BE094404.D
 Acq On : 15 Oct 2017 21:50
 Operator : SJ/JU
 Sample : I5548-17
 Misc :
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 C0AF7

Manual Integrations
 APPROVED

Sohil
 10/16/2017 8:26:41 PM

Quant Time: Oct 16 08:18:33 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE092017.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 16 07:45:00 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.92	152	2073	0.40	ng/ul	0.00
2) Naphthalene-d8	10.70	136	10414	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.53	164	6261	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.26	188	11891	0.40	ng/ul	0.00
16) Chrysene-d12	21.44	240	10687	0.40	ng/ul	0.02
20) Perylene-d12	24.00	264	11384m	0.40	ng/ul	0.05
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.29	152	3221	0.20	ng/ul	0.00
14) Fluoranthene-d10	19.29	212	7272	0.18	ng/ul	0.01
Target Compounds					Ovalue	
3) Naphthalene	10.75	128	22873	0.897	ng/ul#	91
5) 2-Methylnaphthalene	12.36	142	5316	0.321	ng/ul	99
7) Acenaphthylene	14.25	152	35841	1.468	ng/ul	98
8) Acenaphthene	14.59	153	8226	0.409	ng/ul	97
9) Fluorene	15.58	166	15623	0.664	ng/ul	91
12) Phenanthrene	17.31	178	260207	8.374	ng/ul#	94
13) Anthracene	17.39	178	86380	2.858	ng/ul#	93
15) Fluoranthene	19.32	202	618220	15.142	ng/ul	84
17) Pyrene	19.69	202	576917	19.402	ng/ul#	83
18) Benzo(a)anthracene	21.42	228	329711	11.629	ng/ul#	89
19) Chrysene	21.48	228	340346m	10.395	ng/ul	
21) Benzo(b)fluoranthene	23.22	252	450494	14.568	ng/ul	94
22) Benzo(k)fluoranthene	23.24	252	158242m	4.427	ng/ul	
23) Benzo(a)pyrene	23.89	252	360511m	11.219	ng/ul	
24) Indeno(1,2,3-cd)pyrene	26.74	276	257796m	7.842	ng/ul	
25) Dibenzo(a,h)anthracene	26.73	278	65819m	2.387	ng/ul	
26) Benzo(a,h,i)perylene	27.58	276	248364m	8.759	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA E\DATA\BE101417\
Data File : BE094404.D
Acq On : 15 Oct 2017 21:50
Operator : SJ/JU
Sample : I5548-17
Misc :
ALS Vial : 50 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampled :
C0AF7

Manual Integrations
APPROVED

Sohil
10/16/2017 8:26:41 PM

Quant Time: Oct 16 08:18:33 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE092017.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 16 07:45:00 2017
Response via : Initial Calibration

