

Data Path : Z:\HPCHEM1\BNA_E\Data\BE101515\
 Data File : BE091014.D
 Acq On : 15 Oct 2015 9:42
 Operator : UM/IZ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.479

Quant Time: Oct 15 15:39:39 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM02.2-EPA-SIM-BE101415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 15 05:50:36 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
3	Naphthalene	0.929	0.857	7.8	102	0.00
4 SURR	2-Methylnaphthalene-d10	0.446	0.430	3.6	101	0.00
5	2-Methylnaphthalene	0.510	0.484	5.1	103	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
7	Acenaphthylene	1.477	1.412	4.4	103	0.00
8 C	Acenaphthene	1.065	1.019	4.3	102	0.00
9	Fluorene	1.320	1.250	5.3	100	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	101	0.00
11	Pentachlorophenol	0.046	0.031	32.6#	75	0.00
12	Phenanthrene	3.533	3.470	1.8	102	0.00
13	Anthracene	3.464	3.328	3.9	100	0.00
14 SURR	Fluoranthene-d10	0.953	0.897	5.9	100	0.00
15 C	Fluoranthene	4.480	4.233	5.5	100	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
17	Pyrene	3.405	3.279	3.7	99	0.00
18	Benzo(a)anthracene	0.874	0.831	4.9	97	0.00
19	Chrysene	0.934	0.902	3.4	98	0.00
20 I	Perylene-d12	1.000	1.000	0.0	98	0.00
21	Benzo(b)fluoranthene	0.950	0.880	7.4	95	0.00
22	Benzo(k)fluoranthene	0.976	0.973	0.3	100	0.00
23 C	Benzo(a)pyrene	0.932	0.902	3.2	98	0.00
24	Indeno(1,2,3-cd)pyrene	3.724	3.599	3.4	100	0.00
25	Dibenzo(a,h)anthracene	0.940	0.906	3.6	97	0.00
26	Benzo(g,h,i)perylene	3.905	3.845	1.5	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0