

Data Path : Z:\HPCHEM1\BNA_E\Data\BE101515\
 Data File : BE091020.D
 Acq On : 15 Oct 2015 16:13
 Operator : UM/IZ
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.480

Quant Time: Oct 15 18:38:55 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM02.2-EPA-SIM-BE101415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 15 05:50:36 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	-0.01
2 I	Naphthalene-d8	1.000	1.000	0.0	101	-0.01
3	Naphthalene	0.929	0.852	8.3	100	-0.02
4 SURR	2-Methylnaphthalene-d10	0.446	0.428	4.0	99	-0.01
5	2-Methylnaphthalene	0.510	0.497	2.5	104	-0.01
6 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.00
7	Acenaphthylene	1.477	1.438	2.6	102	-0.02
8 C	Acenaphthene	1.065	1.027	3.6	100	-0.02
9	Fluorene	1.320	1.253	5.1	97	-0.02
10 I	Phenanthrene-d10	1.000	1.000	0.0	99	0.00
11	Pentachlorophenol	0.046	0.027	41.3#	64	0.00
12	Phenanthrene	3.533	3.497	1.0	100	0.00
13	Anthracene	3.464	3.290	5.0	97	-0.01
14 SURR	Fluoranthene-d10	0.953	0.868	8.9	94	0.00
15 C	Fluoranthene	4.480	4.115	8.1	94	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	91	0.00
17	Pyrene	3.405	3.356	1.4	93	0.00
18	Benzo(a)anthracene	0.874	0.838	4.1	90	0.00
19	Chrysene	0.934	0.893	4.4	90	0.00
20 I	Perylene-d12	1.000	1.000	0.0	88	0.00
21	Benzo(b)fluoranthene	0.950	0.978	-2.9	95	0.00
22	Benzo(k)fluoranthene	0.976	0.904	7.4	83	0.00
23 C	Benzo(a)pyrene	0.932	0.924	0.9	90	0.00
24	Indeno(1,2,3-cd)pyrene	3.724	3.697	0.7	92	-0.01
25	Dibenzo(a,h)anthracene	0.940	0.928	1.3	89	-0.01
26	Benzo(g,h,i)perylene	3.905	3.934	-0.7	91	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0