

Data Path : Z:\HPCHEM1\BNA_E\Data\BE101617\
 Data File : BE094417.D
 Acq On : 16 Oct 2017 14:50
 Operator : SJ/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE101617

Quant Time: Oct 16 15:50:15 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE101617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 16 14:54:40 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
2	1,4-Dioxane	0.698	0.663	5.0	88	0.00
3	n-Nitrosodimethylamine	0.695	0.667	4.0	89	0.00
4 S	2-Fluorophenol	1.352	1.380	-2.1	94	0.01
5 S	Phenol-d6	2.108	2.145	-1.8	92	0.00
6	bis(2-Chloroethyl)ether	1.559	1.591	-2.1	91	0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	94	0.00
8 S	Nitrobenzene-d5	0.348	0.330	5.2	90	-0.02
9	Naphthalene	4.473	4.548	-1.7	95	0.00
10	Hexachlorobutadiene	0.179	0.177	1.1	93	0.00
11 SURR	2-Methylnaphthalene-d10	0.617	0.633	-2.6	96	0.00
12	2-Methylnaphthalene	0.737	0.755	-2.4	97	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	98	0.00
14 S	2,4,6-Tribromophenol	0.152	0.156	-2.6	94	0.00
15 S	2-Fluorobiphenyl	1.428	1.416	0.8	94	0.00
16	Acenaphthylene	2.169	2.141	1.3	96	0.00
17	Acenaphthene	1.366	1.362	0.3	98	0.00
18	Fluorene	1.761	1.739	1.2	97	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	90	0.00
20	4-Bromophenyl-phenylether	0.200	0.195	2.5	98	0.00
21	Hexachlorobenzene	0.365	0.378	-3.6	96	0.00
22	Pentachlorophenol	0.032	0.030	6.3	92	0.03
23	Phenanthrene	1.574	1.554	1.3	86	0.00
24	Anthracene	1.216	1.274	-4.8	98	0.00
25 SURR	Fluoranthene-d10	6.441	6.670	-3.6	97	0.00
26	Fluoranthene	1.954	1.986	-1.6	97	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	98	0.00
28	Pyrene	1.509	1.457	3.4	96	0.00
29 S	Terphenyl-d14	0.807	0.789	2.2	96	0.00
30	Benzo(a)anthracene	1.384	1.353	2.2	97	-0.01
31	Chrysene	1.332	1.317	1.1	96	-0.01
32	Bis(2-ethylhexyl)phthalate	3.984	3.840	3.6	99	0.00
33	Indeno(1,2,3-cd)pyrene	1.341	1.317	1.8	97	0.00
34 I	Perylene-d12	1.000	1.000	0.0	99	0.00
35	Benzo(b)fluoranthene	1.357	1.307	3.7	95	0.00
36	Benzo(k)fluoranthene	1.337	1.342	-0.4	98	0.00
37 C	Benzo(a)pyrene	1.275	1.261	1.1	98	0.00
38	Dibenzo(a,h)anthracene	1.207	1.206	0.1	98	0.00
39	Benzo(g,h,i)perylene	1.161	1.138	2.0	96	0.00

Data Path : Z:\HPCHEM1\BNA_E\Data\BE101617\
Data File : BE094417.D
Acq On : 16 Oct 2017 14:50
Operator : SJ/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE101617

Quant Time: Oct 16 15:50:15 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE101617.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 16 14:54:40 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		