

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE101716\
 Data File : BE092460.D
 Acq On : 17 Oct 2016 19:44
 Operator : UM/SJ
 Sample : PB93716BS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 PB93716BS

Manual Integrations
 APPROVED

sohil
 10/18/2016 10:59:19 AM

Quant Time: Oct 18 01:02:30 2016
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE092216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Sep 22 18:25:56 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.17	152	14921	5.00	ng	-0.02
7) Naphthalene-d8	10.95	136	65925	5.00	ng	0.00
12) Acenaphthene-d10	14.82	164	41933	5.00	ng	0.00
18) Phenanthrene-d10	17.58	188	72838	5.00	ng	0.02
24) Chrysene-d12	21.75	240	106652	5.00	ng	0.00
31) Perylene-d12	24.11	264	96367	5.00	ng	0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.67	112	26248	7.46	ng	-0.02
5) Phenol-d6	7.34	99	33436	6.67	ng	-0.02
8) Nitrobenzene-d5	9.32	82	16291	3.48	ng	-0.04
13) 2,4,6-Tribromophenol	16.31	330	8606	5.83	ng	0.00
14) 2-Fluorobiphenyl	13.43	172	37935	3.82	ng	-0.01
26) Terphenyl-d14	20.17	244	45003	3.40	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.41	88	4965	3.07	ng	# 84
3) n-Nitrosodimethylamine	3.76	42	8560	3.67	ng	# 95
6) bis(2-Chloroethyl)ether	7.59	93	13716	3.50	ng	# 27
9) Naphthalene	10.99	128	48065	3.37	ng	# 95
10) Hexachlorobutadiene	11.28	225	7225	3.22	ng	100
11) 2-Methylnaphthalene	12.62	142	32446	3.43	ng	# 87
15) Acenaphthylene	14.53	152	209079	3.44	ng	99
16) Acenaphthene	14.87	154	31734	3.34	ng	94
17) Fluorene	15.86	166	41150	3.41	ng	99
19) Hexachlorobenzene	16.89	284	12159m	3.90	ng	
20) Pentachlorophenol	17.23	266	10324	9.03	ng	85
21) Phenanthrene	17.60	178	65994	3.94	ng	# 94
22) Anthracene	17.69	178	67608	4.20	ng	99
23) Fluoranthene	19.61	202	288872	3.63	ng	99
25) Pyrene	19.97	202	313364	3.39	ng	99
27) Benzo(a)anthracene	21.72	228	380904m	3.66	ng	
28) Chrysene	21.77	228	341534m	3.59	ng	
29) Bis(2-ethylhexyl)phthalate	21.63	149	29705	3.32	ng	# 74
30) Indeno(1,2,3-cd)pyrene	26.60	276	407133	4.03	ng	98
32) Benzo(b)fluoranthene	23.38	252	92959	3.98	ng	# 96
33) Benzo(k)fluoranthene	23.43	252	84408	3.46	ng	94
34) Benzo(a)pyrene	24.01	252	87247	3.81	ng	# 95
35) Dibenzo(a,h)anthracene	26.63	278	84610	4.03	ng	# 94
36) Benzo(g,h,i)perylene	27.38	276	350417	3.89	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA_E\DATA\BE101716\
Data File : BE092460.D
Acq On : 17 Oct 2016 19:44
Operator : UM/SJ
Sample : PB93716BS
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
PB93716BS

Manual Integrations
APPROVED

sohil
10/18/2016 10:59:19 AM

Quant Time: Oct 18 01:02:30 2016
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE092216.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Sep 22 18:25:56 2016
Response via : Initial Calibration





#1


1,4-Dichlorobenzene-d4
Concen: 5.00 ng
RT: 8.17 min Scan# 576
Delta R.T. -0.02 min
Lab File: BE092460.D
Acq: 17 Oct 2016 19:44

Instrument :
BNA_E
ClientSampleId :
PB93716BS




Tgt Ion	Ratio	Lower	Upper
152	100		
150	157.3	106.0	159.0
115	57.3	31.2	46.8

Manual Integrations
APPROVED

sohil
10/18/2016 10:59:19 AM




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#2




1,4-Dioxane
Concen: 3.07 ng
RT: 3.41 min Scan# 71
Delta R.T. -0.01 min
Lab File: BE092460.D
Acq: 17 Oct 2016 19:44

Tgt Ion	Ratio	Lower	Upper
88	100		
58	90.9	63.6	95.4
43	48.0	28.0	42.0



Abundance Ion 88.00 (87.70 to 88.70): BE0
Ion 58.00 (57.70 to 58.70): BE0
Ion 43.00 (42.70 to 43.70): BE0









#3
n-Nitrosodimethylamine
Concen: 3.67 ng
RT: 3.76 min Scan# 101
Delta R.T. -0.04 min
Lab File: BE092460.D
Acq: 17 Oct 2016 19:44

Tgt Ion	Ratio	Lower	Upper
42	100		
74	102.4	86.0	129.0
44	8.8	5.1	7.7#

Instrument :
BNA_E
ClientSampleId :
PB93716BS


Manual Integrations
APPROVED

sohil
10/18/2016 10:59:19 AM



#4
2-Fluorophenol
Concen: 7.46 ng
RT: 5.67 min Scan# 363
Delta R.T. -0.02 min
Lab File: BE092460.D
Acq: 17 Oct 2016 19:44

Tgt Ion	Ratio	Lower	Upper
112	100		
64	67.7	53.5	80.3
63	36.4	27.9	41.9





#5

Phenol-d6

Concen: 6.67 ng

RT: 7.34 min Scan# 540

Delta R.T. -0.02 min

Lab File: BE092460.D

Acq: 17 Oct 2016 19:44

Instrument :

BNA_E





ClientSampleId :

PB93716BS

Tgt Ion:	99	Resp:	33436
Ion Ratio	Lower	Upper	
99	100		
42	0.0	16.0	24.0#
71	0.0	30.6	45.8#

Manual Integrations
APPROVED

sohil
10/18/2016 10:59:19 AM



#6

bis(2-Chloroethyl)ether

Concen: 3.50 ng




RT: 7.59 min Scan# 551

Delta R.T. -0.02 min

Lab File: BE092460.D

Acq: 17 Oct 2016 19:44

Tgt Ion:	93	Resp:	13716
Ion Ratio	Lower	Upper	
93	100		
63	0.0	71.6	107.4#
95	32.7	24.0	36.0





#7

Naphthalene-d8

Concen: 5.00 ng

RT: 10.95 min Scan# 708

Delta R.T. 0.00 min

Lab File: BE092460.D

Acq: 17 Oct 2016 19:44

Instrument :

BNA_E





ClientSampleId :

PB93716BS

Tgt Ion	Ratio	Resp	Lower	Upper
136	100	65925		
137	11.8	8.2	12.4	
54	4.5	9.6	14.4	#
68	3.7	6.7	10.1	#

Manual Integrations
APPROVED

sohil
10/18/2016 10:59:19 AM



#8

Nitrobenzene-d5

Concen: 3.48 ng




RT: 9.32 min Scan# 626

Delta R.T. -0.04 min

Lab File: BE092460.D

Acq: 17 Oct 2016 19:44

Tgt Ion	Ratio	Resp	Lower	Upper
82	100	16291		
128	9.5	39.0	58.4	#
54	88.2	44.8	67.2	#





#9

Naphthalene

Concen: 3.37 ng

RT: 10.99 min Scan# 710

Delta R.T. -0.02 min

Lab File: BE092460.D

Acq: 17 Oct 2016 19:44

Instrument :

BNA_E





ClientSampleId :

PB93716BS

Tgt Ion:	128	Resp:	48065
Ion Ratio	Lower	Upper	
128	100		
129	11.1	11.2	16.8#
127	13.1	11.6	17.4

Manual Integrations
APPROVED

sohil
10/18/2016 10:59:19 AM



#10

Hexachlorobutadiene

Concen: 3.22 ng




RT: 11.28 min Scan# 725


Delta R.T. -0.02 min

Lab File: BE092460.D

Acq: 17 Oct 2016 19:44

Tgt Ion:	225	Resp:	7225
Ion Ratio	Lower	Upper	
225	100		
223	63.1	50.6	76.0
227	63.7	51.4	77.0









#11
2-Methylnaphthalene
Concen: 3.43 ng
RT: 12.62 min Scan# 814
Delta R.T. -0.02 min
Lab File: BE092460.D
Acq: 17 Oct 2016 19:44

Instrument :
BNA_E
ClientSampleId :
PB93716BS

Tgt Ion:142 Resp: 32446
Ion Ratio Lower Upper
142 100
141 82.3 73.7 110.5
115 30.5 34.0 51.0#




Manual Integrations
APPROVED


sohil
10/18/2016 10:59:19 AM



#12
Acenaphthene-d10
Concen: 5.00 ng
RT: 14.82 min Scan# 980
Delta R.T. 0.00 min
Lab File: BE092460.D
Acq: 17 Oct 2016 19:44

Tgt Ion:164 Resp: 41933
Ion Ratio Lower Upper
164 100
162 98.8 71.4 107.0
160 42.5 27.4 41.2#









#13
 2,4,6-Tribromophenol
 Concen: 5.83 ng
 RT: 16.31 min Scan# 1088
 Delta R.T. -0.01 min
 Lab File: BE092460.D
 Acq: 17 Oct 2016 19:44

Instrument :
 BNA_E
 ClientSampleId :
 PB93716BS

Tgt Ion	Ratio	Lower	Upper
330	100		
332	96.7	75.4	113.0
141	43.4	31.0	46.6




Manual Integrations
 APPROVED

sohil
 10/18/2016 10:59:19 AM



#14
 2-Fluorobiphenyl
 Concen: 3.82 ng
 RT: 13.43 min Scan# 884
 Delta R.T. -0.01 min
 Lab File: BE092460.D
 Acq: 17 Oct 2016 19:44

Tgt Ion	Ratio	Lower	Upper
172	100		
171	33.8	30.1	45.1
170	21.5	21.8	32.6#





#15

Acenaphthylene

Concen: 3.44 ng

RT: 14.53 min Scan# 963

Delta R.T. 0.00 min

Lab File: BE092460.D

Acq: 17 Oct 2016 19:44

Instrument :

BNA_E





ClientSampleId :

PB93716BS

Tgt Ion:	152	Resp:	209079
Ion Ratio	Lower	Upper	
152	100		
151	4.9	3.9	5.9
153	13.3	10.4	15.6

Manual Integrations
APPROVED

sohil
10/18/2016 10:59:19 AM



#16

Acenaphthene

Concen: 3.34 ng




RT: 14.87 min Scan# 983

Delta R.T. -0.02 min

Lab File: BE092460.D

Acq: 17 Oct 2016 19:44

Tgt Ion:	154	Resp:	31734
Ion Ratio	Lower	Upper	
154	100		
153	451.0	350.8	526.2
152	227.7	171.1	256.7





#17

Fluorene

Concen: 3.41 ng

RT: 15.86 min Scan# 1049

Delta R.T. -0.01 min

Lab File: BE092460.D

Acq: 17 Oct 2016 19:44

Instrument :

BNA_E





ClientSampleId :

PB93716BS

Tgt Ion	Ratio	Lower	Upper
166	100		
165	98.3	78.2	117.4
167	13.4	11.0	16.4

Manual Integrations
APPROVED

sohil
10/18/2016 10:59:19 AM



#18

Phenanthrene-d10

Concen: 5.00 ng




RT: 17.58 min Scan# 1154

Delta R.T. 0.02 min

Lab File: BE092460.D

Acq: 17 Oct 2016 19:44

Tgt Ion	Ratio	Lower	Upper
188	100		
94	2.7	3.2	4.8#
80	2.2	2.6	3.8#









#19
Hexachlorobenzene
Concen: 3.90 ng m
RT: 16.89 min Scan# 1124
Delta R.T. -0.00 min
Lab File: BE092460.D
Acq: 17 Oct 2016 19:44

Instrument :
BNA_E
ClientSampleId :
PB93716BS

Tgt Ion	Ratio	Resp Lower	Resp Upper
284	100		
142	0.0	29.1	43.7#
249	27.8	27.9	41.9#




Manual Integrations
APPROVED


sohil
10/18/2016 10:59:19 AM



#20
Pentachlorophenol
Concen: 9.03 ng
RT: 17.23 min Scan# 1139
Delta R.T. -0.02 min
Lab File: BE092460.D
Acq: 17 Oct 2016 19:44

Tgt Ion	Ratio	Resp Lower	Resp Upper
266	100		
268	62.6	62.5	93.7
264	61.7	57.2	85.8





#21

Phenanthrene

Concen: 3.94 ng

RT: 17.60 min Scan# 1155

Delta R.T. -0.00 min

Lab File: BE092460.D

Acq: 17 Oct 2016 19:44

Instrument :

BNA_E





ClientSampleId :

PB93716BS

Tgt Ion	Ratio	Lower	Upper
178	100		
176	20.2	15.8	23.8
179	15.0	16.0	24.0

Manual Integrations
APPROVED

sohil
10/18/2016 10:59:19 AM



#22

Anthracene

Concen: 4.20 ng




RT: 17.69 min Scan# 1159


Delta R.T. -0.00 min

Lab File: BE092460.D

Acq: 17 Oct 2016 19:44

Tgt Ion	Ratio	Lower	Upper
178	100		
176	19.5	15.0	22.4
179	14.7	12.2	18.2





#23

Fluoranthene

Concen: 3.63 ng

RT: 19.61 min Scan# 1266

Delta R.T. -0.00 min

Lab File: BE092460.D

Acq: 17 Oct 2016 19:44

Instrument :

BNA_E





ClientSampleId :

PB93716BS

Tgt	Ion	202	Resp:	288872
Ion	Ratio	Lower	Upper	
202	100			
101	2.9	2.2	3.4	
203	17.8	13.7	20.5	

Manual Integrations
APPROVED

sohil
10/18/2016 10:59:19 AM



#24

Chrysene-d12

Concen: 5.00 ng




RT: 21.75 min Scan# 1375

Delta R.T. -0.00 min

Lab File: BE092460.D

Acq: 17 Oct 2016 19:44

Tgt	Ion	240	Resp:	106652
Ion	Ratio	Lower	Upper	
240	100			
120	1.6	2.6	4.0#	
236	26.7	24.6	37.0	









#25
 Pyrene
 Concen: 3.39 ng
 RT: 19.97 min Scan# 1287
 Delta R.T. -0.00 min
 Lab File: BE092460.D
 Acq: 17 Oct 2016 19:44

Instrument :
 BNA_E
 ClientSampleId :
 PB93716BS

Tgt Ion	Ratio	Resp Lower	Resp Upper
202	100		
200	5.3	4.2	6.4
203	17.9	13.9	20.9




Manual Integrations
 APPROVED

sohil
 10/18/2016 10:59:19 AM



#26
 Terphenyl-d14
 Concen: 3.40 ng
 RT: 20.17 min Scan# 1299
 Delta R.T. -0.00 min
 Lab File: BE092460.D
 Acq: 17 Oct 2016 19:44

Tgt Ion	Ratio	Resp Lower	Resp Upper
244	100		
212	9.5	6.7	10.1
122	10.4	3.6	5.4#









#27
Benzo(a)anthracene
Concen: 3.66 ng m
RT: 21.72 min Scan# 1374
Delta R.T. -0.00 min
Lab File: BE092460.D
Acq: 17 Oct 2016 19:44

Instrument :
BNA_E
ClientSampleId :
PB93716BS

Tgt Ion	Ratio	Lower	Upper
228	100		
226	23.2	23.6	35.4#
229	19.4	13.3	19.9




Manual Integrations
APPROVED

sohil
10/18/2016 10:59:19 AM



#28
Chrysene
Concen: 3.59 ng m
RT: 21.77 min Scan# 1376
Delta R.T. 0.00 min
Lab File: BE092460.D
Acq: 17 Oct 2016 19:44

Tgt Ion	Ratio	Lower	Upper
228	100		
226	38.7	36.3	54.5
229	12.9	10.6	16.0





#29

Bis(2-ethylhexyl)phthalate

Concen: 3.32 ng

RT: 21.63 min Scan# 1370

Delta R.T. -0.00 min

Lab File: BE092460.D

Acq: 17 Oct 2016 19:44

Instrument :

BNA_E





ClientSampleId :

PB93716BS

Tgt Ion	Ratio	Lower	Upper
149	100		
167	23.6	16.0	24.0
279	0.0	16.0	24.0

Manual Integrations
APPROVED

sohil
10/18/2016 10:59:19 AM



#30

Indeno(1,2,3-cd)pyrene

Concen: 4.03 ng




RT: 26.60 min Scan# 1716

Delta R.T. 0.02 min

Lab File: BE092460.D

Acq: 17 Oct 2016 19:44

Tgt Ion	Ratio	Lower	Upper
276	100		
138	27.1	20.3	30.5
277	24.6	19.7	29.5









#31
 Perylene-d12
 Concen: 5.00 ng
 RT: 24.11 min Scan# 1546
 Delta R.T. 0.02 min
 Lab File: BE092460.D
 Acq: 17 Oct 2016 19:44

Instrument :
 BNA_E
 ClientSampleId :
 PB93716BS

Tgt Ion	Ratio	Lower	Upper
264	100		
260	25.9	20.1	30.1
265	21.0	17.9	26.9




Manual Integrations
 APPROVED

sohil
 10/18/2016 10:59:19 AM



#32
 Benzo(b)fluoranthene
 Concen: 3.98 ng
 RT: 23.38 min Scan# 1483
 Delta R.T. 0.02 min
 Lab File: BE092460.D
 Acq: 17 Oct 2016 19:44

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.7	18.7	28.1
125	9.7	10.5	15.7#





#33

Benzo(k)fluoranthene

Concen: 3.46 ng

RT: 23.43 min Scan# 1487

Delta R.T. 0.02 min

Lab File: BE092460.D

Acq: 17 Oct 2016 19:44

Instrument :

BNA_E





ClientSampleId :

PB93716BS

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.2	20.3	30.5
125	9.9	9.6	14.4

Manual Integrations
APPROVED

sohil
10/18/2016 10:59:19 AM



#34

Benzo(a)pyrene

Concen: 3.81 ng




RT: 24.01 min Scan# 1537


Delta R.T. 0.02 min

Lab File: BE092460.D

Acq: 17 Oct 2016 19:44

Tgt Ion	Ratio	Lower	Upper
252	100		
253	22.8	19.0	28.6
125	10.9	11.8	17.8









#35
Dibenzo(a,h)anthracene
Concen: 4.03 ng
RT: 26.63 min Scan# 1718
Delta R.T. 0.04 min
Lab File: BE092460.D
Acq: 17 Oct 2016 19:44

Instrument :
BNA_E
ClientSampleId :
PB93716BS

Tgt Ion	Ratio	Lower	Upper
278	100		
139	13.8	13.8	20.8#
279	24.3	21.4	32.2

Manual Integrations
APPROVED

sohil
10/18/2016 10:59:19 AM



#36
Benzo(g,h,i)perylene
Concen: 3.89 ng
RT: 27.38 min Scan# 1762
Delta R.T. 0.04 min
Lab File: BE092460.D
Acq: 17 Oct 2016 19:44

Tgt Ion	Ratio	Lower	Upper
276	100		
277	24.5	19.8	29.8
138	22.1	19.8	29.6

