

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE101717\
 Data File : BE094428.D
 Acq On : 17 Oct 2017 12:26
 Operator : SJ/JU
 Sample : PB102812BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 PB102812BL

Quant Time: Oct 17 20:27:14 2017
 Quant Method : Z:\HPCHEM1\BNA_E\Methods\8270-SIM-BE101617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 16 14:54:40 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.90	152	58759	0.40	ng	-0.02
7) Naphthalene-d8	10.69	136	299878	0.40	ng	-0.02
13) Acenaphthene-d10	14.51	164	179355	0.40	ng	-0.01
19) Phenanthrene-d10	17.24	188	375417	0.40	ng	-0.03
27) Chrysene-d12	21.43	240	429699	0.40	ng	0.00
34) Perylene-d12	23.96	264	421554	0.40	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.51	112	562105	2.83	ng	0.00
5) Phenol-d6	7.10	99	838065	2.71	ng	-0.03
8) Nitrobenzene-d5	9.06	82	529224	2.03	ng	-0.02
11) 2-Methylnaphthalene-d10	0.00	152	0d	0.00	ng	
14) 2,4,6-Tribromophenol	16.03	330	148778	2.18	ng	0.00
15) 2-Fluorobiphenyl	13.15	172	1107368	1.73	ng	0.00
25) Fluoranthene-d10	0.00	212	0d	0.00	ng	
29) Terphenyl-d14	19.87	244	1409141	1.63	ng	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE101717\
Data File : BE094428.D
Acq On : 17 Oct 2017 12:26
Operator : SJ/JU
Sample : PB102812BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
PB102812BL

Quant Time: Oct 17 20:27:14 2017
Quant Method : Z:\HPCHEM1\BNA_E\Methods\8270-SIM-BE101617.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 16 14:54:40 2017
Response via : Initial Calibration

