

Data Path : Z:\HPCHEM1\BNA E\DATA\BE101915\
 Data File : BE091022.D
 Acq On : 19 Oct 2015 14:17
 Operator : UM/NP
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.481

Quant Time: Oct 20 00:58:33 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE101415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 15 05:50:36 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	105	0.00
3	Naphthalene	0.929	0.829	10.8	101	0.00
4 SURR	2-Methylnaphthalene-d10	0.446	0.413	7.4	100	0.00
5	2-Methylnaphthalene	0.510	0.488	4.3	106	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	109	0.00
7	Acenaphthylene	1.477	1.382	6.4	106	0.00
8 C	Acenaphthene	1.065	0.970	8.9	102	0.00
9	Fluorene	1.320	1.226	7.1	103	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	110	0.00
11	Pentachlorophenol	0.046	0.025	45.7#	67	0.00
12	Phenanthrene	3.533	3.382	4.3	108	0.00
13	Anthracene	3.464	3.189	7.9	104	0.00
14 SURR	Fluoranthene-d10	0.953	0.842	11.6	102	0.00
15 C	Fluoranthene	4.480	4.048	9.6	103	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	102	0.00
17	Pyrene	3.405	3.229	5.2	101	0.00
18	Benzo(a)anthracene	0.874	0.789	9.7	95	0.00
19	Chrysene	0.934	0.872	6.6	98	0.00
20 I	Perylene-d12	1.000	1.000	0.0	96	0.00
21	Benzo(b)fluoranthene	0.950	0.870	8.4	92	0.00
22	Benzo(k)fluoranthene	0.976	0.916	6.1	93	0.00
23 C	Benzo(a)pyrene	0.932	0.840	9.9	90	0.00
24	Indeno(1,2,3-cd)pyrene	3.724	3.529	5.2	96	0.00
25	Dibenzo(a,h)anthracene	0.940	0.881	6.3	92	0.00
26	Benzo(a,h,i)perylene	3.905	3.703	5.2	94	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0