

Data Path : Z:\HPCHEM1\BNA E\DATA\BE101915\
 Data File : BE091026.D
 Acq On : 19 Oct 2015 18:28
 Operator : UM/NP
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.482

Quant Time: Oct 20 01:02:10 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM02.2-EPA-SIM-BE101415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 20 01:00:24 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	93	0.00
3	Naphthalene	0.929	0.844	9.1	91	0.00
4 SURR	2-Methylnaphthalene-d10	0.446	0.430	3.6	92	0.00
5	2-Methylnaphthalene	0.510	0.483	5.3	94	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	92	0.00
7	Acenaphthylene	1.477	1.461	1.1	95	0.00
8 C	Acenaphthene	1.065	1.010	5.2	90	0.00
9	Fluorene	1.320	1.272	3.6	91	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	91	0.00
11	Pentachlorophenol	0.046	0.020	56.5#	45	-0.01
12	Phenanthrene	3.533	3.482	1.4	93	0.00
13	Anthracene	3.464	3.294	4.9	90	0.00
14 SURR	Fluoranthene-d10	0.953	0.876	8.1	88	0.00
15 C	Fluoranthene	4.480	4.169	6.9	89	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	85	0.00
17	Pyrene	3.405	3.327	2.3	86	0.00
18	Benzo(a)anthracene	0.874	0.805	7.9	80	0.00
19	Chrysene	0.934	0.889	4.8	83	0.00
20 I	Perylene-d12	1.000	1.000	0.0	78	0.00
21	Benzo(b)fluoranthene	0.950	0.964	-1.5	83	0.00
22	Benzo(k)fluoranthene	0.976	0.925	5.2	76	0.00
23 C	Benzo(a)pyrene	0.932	0.889	4.6	77	0.00
24	Indeno(1,2,3-cd)pyrene	3.724	3.627	2.6	80	0.00
25	Dibenzo(a,h)anthracene	0.940	0.910	3.2	78	0.00
26	Benzo(a,h,i)perylene	3.905	3.871	0.9	80	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0