

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE101917\
 Data File : BE094464.D
 Acq On : 20 Oct 2017 2:23
 Operator : SJ/JU
 Sample : I5819-09
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 3910-SPP3

Quant Time: Oct 20 14:00:55 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE101617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 16 14:54:40 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.90	152	1424	0.40	ng	-0.02
7) Naphthalene-d8	10.69	136	7716	0.40	ng	-0.02
13) Acenaphthene-d10	14.51	164	4558	0.40	ng	-0.01
19) Phenanthrene-d10	17.24	188	13004	0.40	ng	-0.03
27) Chrysene-d12	21.42	240	10540	0.40	ng	-0.01
34) Perylene-d12	23.94	264	9897	0.40	ng	-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.52	112	1003	0.21	ng	0.00
5) Phenol-d6	7.14	99	990	0.13	ng	0.00
8) Nitrobenzene-d5	9.06	82	3040	0.45	ng	-0.02
11) 2-Methylnaphthalene-d10	12.28	152	5476	0.46	ng	-0.01
14) 2,4,6-Tribromophenol	16.03	330	575	0.33	ng	0.00
15) 2-Fluorobiphenyl	13.13	172	7940	0.49	ng	-0.01
25) Fluoranthene-d10	19.27	212	56695	0.27	ng	-0.02
29) Terphenyl-d14	19.86	244	11444	0.54	ng	0.00
Target Compounds						
32) Bis(2-ethylhexyl)phthalate	21.29	149	148588	1.415	ng	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE101917\
Data File : BE094464.D
Acq On : 20 Oct 2017 2:23
Operator : SJ/JU
Sample : I5819-09
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
3910-SPP3

Quant Time: Oct 20 14:00:55 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE101617.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 16 14:54:40 2017
Response via : Initial Calibration

