

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE101917\
 Data File : BE094441.D
 Acq On : 19 Oct 2017 11:01
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 20 13:02:23 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE101617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 16 14:54:40 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	87	-0.01
2	1,4-Dioxane	0.400	0.468	-17.0	97	0.00
3	n-Nitrosodimethylamine	0.400	0.364	9.0	80	0.01
4 S	2-Fluorophenol	0.400	0.421	-5.2	91	0.00
5 S	Phenol-d6	0.400	0.416	-4.0	88	0.00
6	bis(2-Chloroethyl)ether	0.400	0.434	-8.5	91	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	89	-0.02
8 S	Nitrobenzene-d5	0.400	0.411	-2.7	92	0.00
9	Naphthalene	0.400	0.435	-8.7	96	-0.02
10	Hexachlorobutadiene	0.400	0.409	-2.2	91	-0.02
11 SURR	2-Methylnaphthalene-d10	0.400	0.442	-10.5	98	0.00
12	2-Methylnaphthalene	0.400	0.447	-11.7	100	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	97	0.00
14 S	2,4,6-Tribromophenol	0.400	0.365	8.8	83	0.00
15 S	2-Fluorobiphenyl	0.400	0.428	-7.0	102	0.00
16	Acenaphthylene	0.400	0.421	-5.2	102	0.00
17	Acenaphthene	0.400	0.427	-6.7	104	0.00
18	Fluorene	0.400	0.426	-6.5	104	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	93	-0.03
20	4-Bromophenyl-phenylether	0.400	0.465	-16.3	120	-0.03
21	Hexachlorobenzene	0.400	0.342	14.5	87	0.00
22	Pentachlorophenol	0.400	0.345	13.8	63	0.03
23	Phenanthrene	0.400	0.335	16.3	75	0.00
24	Anthracene	0.400	0.434	-8.5	104	0.00
25 SURR	Fluoranthene-d10	0.400	0.399	0.3	97	0.00
26	Fluoranthene	0.400	0.395	1.3	97	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	96	0.00
28	Pyrene	0.400	0.399	0.3	96	0.00
29 S	Terphenyl-d14	0.400	0.407	-1.7	97	0.00
30	Benzo(a)anthracene	0.400	0.412	-3.0	100	-0.01
31	Chrysene	0.400	0.417	-4.2	98	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.371	7.3	92	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.414	-3.5	100	-0.02
34 I	Perylene-d12	0.400	0.400	0.0	98	0.00
35	Benzo(b)fluoranthene	0.400	0.416	-4.0	101	0.00
36	Benzo(k)fluoranthene	0.400	0.417	-4.2	100	0.00
37 C	Benzo(a)pyrene	0.400	0.414	-3.5	101	0.00
38	Dibenzo(a,h)anthracene	0.400	0.414	-3.5	100	-0.02
39	Benzo(g,h,i)perylene	0.400	0.411	-2.7	99	-0.02

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE101917\
Data File : BE094441.D
Acq On : 19 Oct 2017 11:01
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 20 13:02:23 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE101617.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 16 14:54:40 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				