

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE101917\
 Data File : BE094455.D
 Acq On : 19 Oct 2017 19:40
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 20 12:42:36 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE101617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 16 14:54:40 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	-0.01
2	1,4-Dioxane	0.698	0.803	-15.0	101	0.00
3	n-Nitrosodimethylamine	0.695	0.714	-2.7	91	0.01
4 S	2-Fluorophenol	1.352	1.521	-12.5	98	0.00
5 S	Phenol-d6	2.108	2.318	-10.0	94	0.00
6	bis(2-Chloroethyl)ether	1.559	1.667	-6.9	91	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	90	-0.02
8 S	Nitrobenzene-d5	0.348	0.364	-4.6	95	0.00
9	Naphthalene	4.473	4.833	-8.0	97	-0.02
10	Hexachlorobutadiene	0.179	0.182	-1.7	92	-0.02
11 SURR	2-Methylnaphthalene-d10	0.617	0.686	-11.2	100	0.00
12	2-Methylnaphthalene	0.737	0.833	-13.0	103	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	101	-0.01
14 S	2,4,6-Tribromophenol	0.152	0.162	-6.6	101	0.00
15 S	2-Fluorobiphenyl	1.428	1.470	-2.9	101	0.00
16	Acenaphthylene	2.169	2.271	-4.7	105	0.00
17	Acenaphthene	1.366	1.411	-3.3	104	0.00
18	Fluorene	1.761	1.844	-4.7	106	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	105	-0.03
20	4-Bromophenyl-phenylether	0.200	0.242	-21.0	141	-0.03
21	Hexachlorobenzene	0.365	0.245	32.9#	72	0.00
22	Pentachlorophenol	0.032	0.039	-21.9	137	0.00
23	Phenanthrene	1.574	1.126	28.5#	72	0.00
24	Anthracene	1.216	1.206	0.8	107	-0.03
25 SURR	Fluoranthene-d10	6.441	5.923	8.0	100	0.00
26	Fluoranthene	1.954	1.790	8.4	101	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
28	Pyrene	1.509	1.498	0.7	100	0.00
29 S	Terphenyl-d14	0.807	0.814	-0.9	100	0.00
30	Benzo(a)anthracene	1.384	1.407	-1.7	103	-0.01
31	Chrysene	1.332	1.332	0.0	98	-0.01
32	Bis(2-ethylhexyl)phthalate	3.984	3.661	8.1	96	0.00
33	Indeno(1,2,3-cd)pyrene	1.341	1.360	-1.4	102	-0.03
34 I	Perylene-d12	1.000	1.000	0.0	100	-0.02
35	Benzo(b)fluoranthene	1.357	1.396	-2.9	103	-0.01
36	Benzo(k)fluoranthene	1.337	1.392	-4.1	103	-0.01
37 C	Benzo(a)pyrene	1.275	1.313	-3.0	103	-0.01
38	Dibenzo(a,h)anthracene	1.207	1.242	-2.9	102	-0.03
39	Benzo(g,h,i)perylene	1.161	1.181	-1.7	101	-0.04

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE101917\
Data File : BE094455.D
Acq On : 19 Oct 2017 19:40
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 20 12:42:36 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE101617.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 16 14:54:40 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		