

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE101917\
 Data File : BE094457.D
 Acq On : 19 Oct 2017 21:30
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 20 13:52:26 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE101617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 16 14:54:40 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	-0.01
2	1,4-Dioxane	0.698	0.715	-2.4	96	0.00
3	n-Nitrosodimethylamine	0.695	0.669	3.7	90	0.01
4 S	2-Fluorophenol	1.352	1.444	-6.8	99	0.00
5 S	Phenol-d6	2.108	2.211	-4.9	95	0.00
6	bis(2-Chloroethyl)ether	1.559	1.664	-6.7	96	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	91	-0.02
8 S	Nitrobenzene-d5	0.348	0.359	-3.2	95	0.00
9	Naphthalene	4.473	4.895	-9.4	100	-0.02
10	Hexachlorobutadiene	0.179	0.186	-3.9	96	-0.02
11 SURR	2-Methylnaphthalene-d10	0.617	0.687	-11.3	101	0.00
12	2-Methylnaphthalene	0.737	0.829	-12.5	104	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	100	-0.01
14 S	2,4,6-Tribromophenol	0.152	0.150	1.3	92	0.00
15 S	2-Fluorobiphenyl	1.428	1.523	-6.7	104	0.00
16	Acenaphthylene	2.169	2.277	-5.0	104	0.00
17	Acenaphthene	1.366	1.433	-4.9	105	0.00
18	Fluorene	1.761	1.834	-4.1	104	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	101	-0.03
20	4-Bromophenyl-phenylether	0.200	0.230	-15.0	130	-0.03
21	Hexachlorobenzene	0.365	0.284	22.2	81	0.00
22	Pentachlorophenol	0.032	0.033	-3.1	113	0.00
23	Phenanthrene	1.574	1.211	23.1	75	0.00
24	Anthracene	1.216	1.237	-1.7	107	-0.03
25 SURR	Fluoranthene-d10	6.441	6.008	6.7	98	0.00
26	Fluoranthene	1.954	1.818	7.0	99	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
28	Pyrene	1.509	1.479	2.0	98	0.00
29 S	Terphenyl-d14	0.807	0.801	0.7	98	0.00
30	Benzo(a)anthracene	1.384	1.412	-2.0	102	-0.01
31	Chrysene	1.332	1.366	-2.6	100	-0.01
32	Bis(2-ethylhexyl)phthalate	3.984	3.537	11.2	91	0.00
33	Indeno(1,2,3-cd)pyrene	1.341	1.362	-1.6	101	-0.03
34 I	Perylene-d12	1.000	1.000	0.0	99	-0.01
35	Benzo(b)fluoranthene	1.357	1.393	-2.7	102	-0.01
36	Benzo(k)fluoranthene	1.337	1.401	-4.8	103	-0.01
37 C	Benzo(a)pyrene	1.275	1.311	-2.8	102	-0.01
38	Dibenzo(a,h)anthracene	1.207	1.254	-3.9	102	-0.03
39	Benzo(g,h,i)perylene	1.161	1.207	-4.0	102	-0.03

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE101917\
Data File : BE094457.D
Acq On : 19 Oct 2017 21:30
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 20 13:52:26 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE101617.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 16 14:54:40 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		