

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE101917\
 Data File : BE094465.D
 Acq On : 20 Oct 2017 3:01
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4EC

Quant Time: Oct 20 14:02:04 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE101617.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 16 14:54:40 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	-0.02
2	1,4-Dioxane	0.698	0.692	0.9	89	0.00
3	n-Nitrosodimethylamine	0.695	0.732	-5.3	95	0.00
4 S	2-Fluorophenol	1.352	1.485	-9.8	98	0.00
5 S	Phenol-d6	2.108	2.311	-9.6	96	0.00
6	bis(2-Chloroethyl)ether	1.559	1.692	-8.5	95	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	92	-0.02
8 S	Nitrobenzene-d5	0.348	0.358	-2.9	95	0.00
9	Naphthalene	4.473	4.818	-7.7	99	-0.02
10	Hexachlorobutadiene	0.179	0.181	-1.1	94	-0.02
11 SURR	2-Methylnaphthalene-d10	0.617	0.678	-9.9	101	0.00
12	2-Methylnaphthalene	0.737	0.818	-11.0	103	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	99	-0.01
14 S	2,4,6-Tribromophenol	0.152	0.162	-6.6	98	0.00
15 S	2-Fluorobiphenyl	1.428	1.488	-4.2	100	0.00
16	Acenaphthylene	2.169	2.247	-3.6	102	0.00
17	Acenaphthene	1.366	1.444	-5.7	105	0.00
18	Fluorene	1.761	1.823	-3.5	102	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	102	-0.03
20	4-Bromophenyl-phenylether	0.200	0.229	-14.5	130	-0.03
21	Hexachlorobenzene	0.365	0.256	29.9#	74	0.00
22	Pentachlorophenol	0.032	0.033	-3.1	115	0.00
23	Phenanthrene	1.574	1.130	28.2#	71	0.00
24	Anthracene	1.216	1.247	-2.5	108	-0.03
25 SURR	Fluoranthene-d10	6.441	6.043	6.2	100	0.00
26	Fluoranthene	1.954	1.826	6.6	100	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	98	0.00
28	Pyrene	1.509	1.536	-1.8	100	0.00
29 S	Terphenyl-d14	0.807	0.833	-3.2	100	0.00
30	Benzo(a)anthracene	1.384	1.407	-1.7	100	-0.01
31	Chrysene	1.332	1.427	-7.1	103	-0.01
32	Bis(2-ethylhexyl)phthalate	3.984	3.664	8.0	93	0.00
33	Indeno(1,2,3-cd)pyrene	1.341	1.383	-3.1	101	-0.03
34 I	Perylene-d12	1.000	1.000	0.0	100	-0.02
35	Benzo(b)fluoranthene	1.357	1.345	0.9	99	-0.01
36	Benzo(k)fluoranthene	1.337	1.417	-6.0	105	-0.01
37 C	Benzo(a)pyrene	1.275	1.305	-2.4	102	-0.02
38	Dibenzo(a,h)anthracene	1.207	1.243	-3.0	102	-0.02
39	Benzo(g,h,i)perylene	1.161	1.188	-2.3	101	-0.03

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE101917\
Data File : BE094465.D
Acq On : 20 Oct 2017 3:01
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4EC

Quant Time: Oct 20 14:02:04 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE101617.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 16 14:54:40 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		