

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE101918\
 Data File : BE097985.D
 Acq On : 19 Oct 2018 9:44
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 19 23:33:41 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 16 15:36:44 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	135	0.00
2	1,4-Dioxane	0.697	0.572	17.9	120	0.00
3	n-Nitrosodimethylamine	0.779	0.707	9.2	138	0.00
4 S	2-Fluorophenol	1.261	1.187	5.9	134	-0.01
5 S	Phenol-d6	1.864	1.855	0.5	138	-0.01
6	bis(2-Chloroethyl)ether	1.583	1.385	12.5	126	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	151#	-0.01
8 S	Nitrobenzene-d5	0.379	0.332	12.4	135	-0.01
9	Naphthalene	3.983	3.792	4.8	148	-0.01
10	Hexachlorobutadiene	0.215	0.201	6.5	146	-0.01
11 SURR	2-Methylnaphthalene-d10	0.665	0.645	3.0	150#	-0.01
12	2-Methylnaphthalene	0.682	0.671	1.6	154#	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	164#	-0.01
14 S	2,4,6-Tribromophenol	0.215	0.180	16.3	156#	-0.01
15 S	2-Fluorobiphenyl	1.635	1.412	13.6	150	-0.01
16	Acenaphthylene	1.848	1.706	7.7	160#	0.00
17	Acenaphthene	1.129	1.064	5.8	163#	0.00
18	Fluorene	1.554	1.377	11.4	153#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	159#	-0.01
20	4-Bromophenyl-phenylether	0.219	0.200	8.7	149	-0.01
21	Hexachlorobenzene	0.215	0.210	2.3	158#	-0.01
22	Pentachlorophenol	0.062	0.063	-1.6	214#	-0.01
23	Phenanthrene	1.016	0.903	11.1	146	-0.01
24	Anthracene	0.960	0.850	11.5	147	-0.01
25 SURR	Fluoranthene-d10	5.421	4.427	18.3	139	-0.01
26	Fluoranthene	1.334	1.107	17.0	141	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	148	-0.01
28	Pvrene	1.154	1.027	11.0	140	0.00
29 S	Terphenyl-d14	0.917	0.791	13.7	141	-0.01
30	Benzo(a)anthracene	1.227	1.079	12.1	138	0.00
31	Chrysene	1.162	1.046	10.0	146	-0.01
32	Bis(2-ethylhexyl)phthalate	2.865	2.504	12.6	132	-0.01
33	Indeno(1,2,3-cd)pyrene	1.249	1.114	10.8	133	-0.03
34 I	Perylene-d12	1.000	1.000	0.0	144	-0.02
35	Benzo(b)fluoranthene	1.151	1.008	12.4	136	-0.01
36	Benzo(k)fluoranthene	1.199	1.059	11.7	141	-0.01
37 C	Benzo(a)pyrene	1.122	0.995	11.3	136	-0.02
38	Dibenzo(a,h)anthracene	1.099	0.955	13.1	134	-0.02
39	Benzo(a,h,i)perylene	1.126	0.950	15.6	133	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE101918\
Data File : BE097985.D
Acq On : 19 Oct 2018 9:44
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 19 23:33:41 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 16 15:36:44 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			