

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE101918\
 Data File : BE097989.D
 Acq On : 19 Oct 2018 12:28
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 19 23:39:30 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 16 15:36:44 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	137	0.00
2	1,4-Dioxane	0.697	0.661	5.2	141	0.00
3	n-Nitrosodimethylamine	0.779	0.690	11.4	138	0.00
4 S	2-Fluorophenol	1.261	1.208	4.2	139	0.00
5 S	Phenol-d6	1.864	1.980	-6.2	150#	0.00
6	bis(2-Chloroethyl)ether	1.583	1.369	13.5	127	0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	159#	0.00
8 S	Nitrobenzene-d5	0.379	0.342	9.8	147	0.00
9	Naphthalene	3.983	3.694	7.3	152#	0.00
10	Hexachlorobutadiene	0.215	0.196	8.8	150#	0.00
11 SURR	2-Methylnaphthalene-d10	0.665	0.654	1.7	160#	0.00
12	2-Methylnaphthalene	0.682	0.659	3.4	159#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	171#	0.00
14 S	2,4,6-Tribromophenol	0.215	0.171	20.5	154#	0.00
15 S	2-Fluorobiphenyl	1.635	1.409	13.8	156#	0.00
16	Acenaphthylene	1.848	1.672	9.5	163#	0.00
17	Acenaphthene	1.129	1.024	9.3	163#	0.00
18	Fluorene	1.554	1.334	14.2	154#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	163#	0.00
20	4-Bromophenyl-phenylether	0.219	0.199	9.1	152#	0.00
21	Hexachlorobenzene	0.215	0.203	5.6	157#	0.00
22	Pentachlorophenol	0.062	0.059	4.8	206#	0.00
23	Phenanthrene	1.016	0.907	10.7	150#	0.00
24	Anthracene	0.960	0.840	12.5	149	0.00
25 SURR	Fluoranthene-d10	5.421	4.302	20.6	139	0.00
26	Fluoranthene	1.334	1.059	20.6	138	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	145	-0.01
28	Pvrene	1.154	1.014	12.1	136	0.00
29 S	Terphenyl-d14	0.917	0.790	13.8	138	0.00
30	Benzo(a)anthracene	1.227	1.076	12.3	136	0.00
31	Chrysene	1.162	1.068	8.1	146	-0.01
32	Bis(2-ethylhexyl)phthalate	2.865	2.393	16.5	124	0.00
33	Indeno(1,2,3-cd)pyrene	1.249	1.126	9.8	132	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	143	-0.01
35	Benzo(b)fluoranthene	1.151	1.023	11.1	136	0.00
36	Benzo(k)fluoranthene	1.199	1.062	11.4	140	0.00
37 C	Benzo(a)pyrene	1.122	0.981	12.6	132	-0.01
38	Dibenzo(a,h)anthracene	1.099	0.966	12.1	134	-0.01
39	Benzo(a,h,i)perylene	1.126	0.961	14.7	133	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE101918\
Data File : BE097989.D
Acq On : 19 Oct 2018 12:28
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 19 23:39:30 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 16 15:36:44 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			