

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE101918\
 Data File : BE097989.D
 Acq On : 19 Oct 2018 12:28
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 19 23:39:30 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 16 15:36:44 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	137	0.00
2	1,4-Dioxane	0.400	0.379	5.3	141	0.00
3	n-Nitrosodimethylamine	0.400	0.354	11.5	138	0.00
4 S	2-Fluorophenol	0.400	0.383	4.3	139	0.00
5 S	Phenol-d6	0.400	0.425	-6.2	150	0.00
6	bis(2-Chloroethyl)ether	0.400	0.346	13.5	127	0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	159	0.00
8 S	Nitrobenzene-d5	0.400	0.362	9.5	147	0.00
9	Naphthalene	0.400	0.371	7.3	152	0.00
10	Hexachlorobutadiene	0.400	0.366	8.5	150	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.393	1.8	160	0.00
12	2-Methylnaphthalene	0.400	0.386	3.5	159	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	171	0.00
14 S	2,4,6-Tribromophenol	0.400	0.318	20.5	154	0.00
15 S	2-Fluorobiphenyl	0.400	0.345	13.8	156	0.00
16	Acenaphthylene	0.400	0.362	9.5	163	0.00
17	Acenaphthene	0.400	0.363	9.3	163	0.00
18	Fluorene	0.400	0.343	14.2	154	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	163	0.00
20	4-Bromophenyl-phenylether	0.400	0.365	8.8	152	0.00
21	Hexachlorobenzene	0.400	0.378	5.5	157	0.00
22	Pentachlorophenol	0.400	0.448	-12.0	206	0.00
23	Phenanthrene	0.400	0.357	10.8	150	0.00
24	Anthracene	0.400	0.350	12.5	149	0.00
25 SURR	Fluoranthene-d10	0.400	0.317	20.8	139	0.00
26	Fluoranthene	0.400	0.318	20.5	138	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	145	-0.01
28	Pvrene	0.400	0.351	12.3	136	0.00
29 S	Terphenyl-d14	0.400	0.345	13.8	138	0.00
30	Benzo(a)anthracene	0.400	0.351	12.3	136	0.00
31	Chrysene	0.400	0.368	8.0	146	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.310	22.5	124	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.331	17.3	132	-0.02
34 I	Perylene-d12	0.400	0.400	0.0	143	-0.01
35	Benzo(b)fluoranthene	0.400	0.355	11.3	136	0.00
36	Benzo(k)fluoranthene	0.400	0.354	11.5	140	0.00
37 C	Benzo(a)pyrene	0.400	0.350	12.5	132	-0.01
38	Dibenzo(a,h)anthracene	0.400	0.352	12.0	134	-0.01
39	Benzo(a,h,i)perylene	0.400	0.342	14.5	133	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE101918\
Data File : BE097989.D
Acq On : 19 Oct 2018 12:28
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 19 23:39:30 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 16 15:36:44 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				