

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE101918\
 Data File : BE097991.D
 Acq On : 19 Oct 2018 14:48
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 19 23:49:11 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 16 15:36:44 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	135	0.00
2	1,4-Dioxane	0.697	0.624	10.5	131	0.00
3	n-Nitrosodimethylamine	0.779	0.637	18.2	125	0.00
4 S	2-Fluorophenol	1.261	1.226	2.8	139	0.00
5 S	Phenol-d6	1.864	1.913	-2.6	142	0.00
6	bis(2-Chloroethyl)ether	1.583	1.433	9.5	130	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	155#	0.00
8 S	Nitrobenzene-d5	0.379	0.326	14.0	136	0.00
9	Naphthalene	3.983	3.774	5.2	151#	0.00
10	Hexachlorobutadiene	0.215	0.198	7.9	148	0.00
11 SURR	2-Methylnaphthalene-d10	0.665	0.651	2.1	155#	0.00
12	2-Methylnaphthalene	0.682	0.678	0.6	159#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	169#	0.00
14 S	2,4,6-Tribromophenol	0.215	0.174	19.1	154#	0.00
15 S	2-Fluorobiphenyl	1.635	1.392	14.9	152#	0.00
16	Acenaphthylene	1.848	1.672	9.5	161#	0.00
17	Acenaphthene	1.129	1.039	8.0	163#	0.00
18	Fluorene	1.554	1.361	12.4	155#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	159#	0.00
20	4-Bromophenyl-phenylether	0.219	0.207	5.5	154#	0.00
21	Hexachlorobenzene	0.215	0.210	2.3	159#	0.00
22	Pentachlorophenol	0.062	0.063	-1.6	214#	-0.01
23	Phenanthrene	1.016	0.909	10.5	147	0.00
24	Anthracene	0.960	0.858	10.6	149	0.00
25 SURR	Fluoranthene-d10	5.421	4.404	18.8	139	0.00
26	Fluoranthene	1.334	1.090	18.3	139	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	145	-0.01
28	Pvrene	1.154	1.032	10.6	138	0.00
29 S	Terphenyl-d14	0.917	0.782	14.7	137	0.00
30	Benzo(a)anthracene	1.227	1.084	11.7	137	0.00
31	Chrysene	1.162	1.064	8.4	146	-0.01
32	Bis(2-ethylhexyl)phthalate	2.865	2.540	11.3	132	0.00
33	Indeno(1,2,3-cd)pyrene	1.249	1.117	10.6	131	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	143	-0.01
35	Benzo(b)fluoranthene	1.151	1.025	10.9	137	0.00
36	Benzo(k)fluoranthene	1.199	1.057	11.8	139	0.00
37 C	Benzo(a)pyrene	1.122	0.983	12.4	133	-0.01
38	Dibenzo(a,h)anthracene	1.099	0.958	12.8	133	0.00
39	Benzo(a,h,i)perylene	1.126	0.963	14.5	134	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA E\Data\BE101918\
Data File : BE097991.D
Acq On : 19 Oct 2018 14:48
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Oct 19 23:49:11 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE101618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 16 15:36:44 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			